

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX011524\
 Data File : VX039748.D
 Acq On : 15 Jan 2024 11:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 16 01:36:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	50.000	46.775	6.5	86	0.00
3 P	Chloromethane	50.000	43.840	12.3	85	0.00
4 C	Vinyl Chloride	50.000	45.100	9.8#	87	0.00
5 T	Bromomethane	50.000	42.193	15.6	81	0.00
6 T	Chloroethane	50.000	47.229	5.5	86	0.00
7 T	Trichlorofluoromethane	50.000	47.278	5.4	92	0.00
8 T	Diethyl Ether	50.000	50.782	-1.6	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.032	3.9	93	0.00
10 T	Methyl Iodide	50.000	56.585	-13.2	101	0.00
11 T	Tert butyl alcohol	250.000	261.875	-4.8	100	0.02
12 CM	1,1-Dichloroethene	50.000	47.101	5.8#	90	0.00
13 T	Acrolein	250.000	267.081	-6.8	100	0.00
14 T	Allyl chloride	50.000	47.438	5.1	90	0.00
15 T	Acrylonitrile	250.000	243.034	2.8	91	0.00
16 T	Acetone	250.000	247.188	1.1	93	0.00
17 T	Carbon Disulfide	50.000	43.122	13.8	84	0.00
18 T	Methyl Acetate	50.000	49.187	1.6	93	0.00
19 T	Methyl tert-butyl Ether	50.000	50.772	-1.5	98	0.00
20 T	Methylene Chloride	50.000	48.474	3.1	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.809	6.4	90	0.00
22 T	Diisopropyl ether	50.000	50.312	-0.6	97	0.00
23 T	Vinyl Acetate	250.000	253.438	-1.4	94	0.00
24 P	1,1-Dichloroethane	50.000	49.184	1.6	93	0.00
25 T	2-Butanone	250.000	238.942	4.4	90	0.00
26 T	2,2-Dichloropropane	50.000	49.882	0.2	95	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.918	0.2	94	0.00
28 T	Bromochloromethane	50.000	50.862	-1.7	96	0.00
29 T	Tetrahydrofuran	250.000	237.648	4.9	90	-0.01
30 C	Chloroform	50.000	50.742	-1.5#	98	0.00
31 T	Cyclohexane	50.000	46.557	6.9	89	0.00
32 T	1,1,1-Trichloroethane	50.000	49.511	1.0	93	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.253	5.5	94	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	93	0.00
35 S	Dibromofluoromethane	50.000	50.445	-0.9	97	0.00
36 T	1,1-Dichloropropene	50.000	47.393	5.2	91	0.00
37 T	Ethyl Acetate	50.000	48.269	3.5	93	0.00
38 T	Carbon Tetrachloride	50.000	48.676	2.6	89	0.00
39 T	Methylcyclohexane	50.000	47.710	4.6	89	0.00
40 TM	Benzene	50.000	49.955	0.1	94	0.00
41 T	Methacrylonitrile	50.000	52.283	-4.6	91	0.00
42 TM	1,2-Dichloroethane	50.000	50.999	-2.0	95	0.00
43 T	Isopropyl Acetate	50.000	51.591	-3.2	93	0.00
44 TM	Trichloroethene	50.000	50.854	-1.7	94	0.00
45 C	1,2-Dichloropropane	50.000	51.434	-2.9#	98	0.00
46 T	Dibromomethane	50.000	51.985	-4.0	96	0.00
47 T	Bromodichloromethane	50.000	52.305	-4.6	96	0.00
48 T	Methyl methacrylate	50.000	50.518	-1.0	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX011524\
 Data File : VX039748.D
 Acq On : 15 Jan 2024 11:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 16 01:36:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1038.216	-3.8	94	0.00
50 S	Toluene-d8	50.000	48.730	2.5	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	253.524	-1.4	93	0.00
52 CM	Toluene	50.000	51.409	-2.8#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	53.974	-7.9	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.278	-6.6	95	0.00
55 T	1,1,2-Trichloroethane	50.000	52.327	-4.7	98	0.00
56 T	Ethyl methacrylate	50.000	53.719	-7.4	97	0.00
57 T	1,3-Dichloropropane	50.000	52.074	-4.1	97	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	268.251	-7.3	96	0.00
59 T	2-Hexanone	250.000	266.475	-6.6	95	0.00
60 T	Dibromochloromethane	50.000	55.911	-11.8	100	0.00
61 T	1,2-Dibromoethane	50.000	54.250	-8.5	99	0.00
62 S	4-Bromofluorobenzene	50.000	49.997	0.0	95	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	91	0.00
64 T	Tetrachloroethene	50.000	49.187	1.6	95	0.00
65 PM	Chlorobenzene	50.000	51.193	-2.4	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.955	-11.9	98	0.00
67 C	Ethyl Benzene	50.000	51.312	-2.6#	92	0.00
68 T	m/p-Xylenes	100.000	104.519	-4.5	95	0.00
69 T	o-Xylene	50.000	51.714	-3.4	95	0.00
70 T	Styrene	50.000	52.738	-5.5	95	0.00
71 P	Bromoform	50.000	58.052	-16.1	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	92	0.00
73 T	Isopropylbenzene	50.000	49.993	0.0	94	0.00
74 T	N-ethyl acetate	50.000	54.104	-8.2	96	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.997	-4.0	98	0.00
76 T	1,2,3-Trichloropropane	50.000	52.358	-4.7	99	0.00
77 T	Bromobenzene	50.000	52.144	-4.3	100	0.00
78 T	n-propylbenzene	50.000	50.020	-0.0	92	0.00
79 T	2-Chlorotoluene	50.000	48.575	2.8	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.189	-2.4	95	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	53.690	-7.4	98	0.00
82 T	4-Chlorotoluene	50.000	49.981	0.0	95	0.00
83 T	tert-Butylbenzene	50.000	51.771	-3.5	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.091	-4.2	96	0.00
85 T	sec-Butylbenzene	50.000	50.893	-1.8	94	0.00
86 T	p-Isopropyltoluene	50.000	51.759	-3.5	94	0.00
87 T	1,3-Dichlorobenzene	50.000	51.281	-2.6	98	0.00
88 T	1,4-Dichlorobenzene	50.000	49.810	0.4	96	0.00
89 T	n-Butylbenzene	50.000	48.934	2.1	88	0.00
90 T	Hexachloroethane	50.000	53.124	-6.2	94	0.00
91 T	1,2-Dichlorobenzene	50.000	50.757	-1.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.008	-2.0	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.438	3.1	92	0.00
94 T	Hexachlorobutadiene	50.000	48.976	2.0	91	0.00
95 T	Naphthalene	50.000	51.090	-2.2	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011524\
Data File : VX039748.D
Acq On : 15 Jan 2024 11:41
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 16 01:36:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.577	0.8	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6