

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
 Data File : VX044652.D
 Acq On : 15 Jan 2025 13:14
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: Jan 16 01:01:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 00:57:36 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 01/15/2025
 Supervised By : Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	170573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	308450	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	255886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	110019	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	235193	91.349	ug/l	-0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	182.700%#
35) Dibromofluoromethane	5.373	113	181010	92.374	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	184.740%#
50) Toluene-d8	8.641	98	617688	90.404	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	180.800%#
62) 4-Bromofluorobenzene	11.079	95	236119	92.702	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	185.400%#
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	256287	111.657	ug/l	99
3) Chloromethane	1.301	50	232745	93.768	ug/l	100
4) Vinyl Chloride	1.374	62	238376	97.198	ug/l	100
5) Bromomethane	1.593	94	57815	98.728	ug/l	94
6) Chloroethane	1.660	64	42244	93.931	ug/l	97
7) Trichlorofluoromethane	1.861	101	275756	101.736	ug/l	99
8) Diethyl Ether	2.130	74	99765	91.205	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.313	101	203976	100.109	ug/l	99
10) Methyl Iodide	2.441	142	272297	97.884	ug/l	99
11) Tert butyl alcohol	3.026	59	259620m	447.152	ug/l	
12) 1,1-Dichloroethene	2.306	96	194870	97.570	ug/l	98
13) Acrolein	2.233	56	116830	432.534	ug/l	98
14) Allyl chloride	2.654	41	358388	92.162	ug/l	100
15) Acrylonitrile	3.062	53	560841	498.179	ug/l	100
16) Acetone	2.386	43	527741	518.081	ug/l	100
17) Carbon Disulfide	2.495	76	529418	97.144	ug/l	99
18) Methyl Acetate	2.697	43	321291	102.235	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	681776	94.245	ug/l	99
20) Methylene Chloride	2.776	84	216105	91.100	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	190588	91.400	ug/l	99
22) Diisopropyl ether	3.751	45	632562	93.579	ug/l	100
23) Vinyl Acetate	3.715	43	2791754	471.863	ug/l	100
24) 1,1-Dichloroethane	3.599	63	378924	94.530	ug/l	99
25) 2-Butanone	4.550	43	765500	505.435	ug/l	100
26) 2,2-Dichloropropane	4.458	77	357666	92.064	ug/l	99
27) cis-1,2-Dichloroethene	4.471	96	240149	91.559	ug/l	99
28) Bromochloromethane	4.885	49	179896	96.041	ug/l	100
29) Tetrahydrofuran	4.995	42	477227	488.500	ug/l	100
30) Chloroform	5.080	83	373681	92.373	ug/l	99
31) Cyclohexane	5.452	56	319131	98.436	ug/l	98
32) 1,1,1-Trichloroethane	5.367	97	338119	90.918	ug/l	99
36) 1,1-Dichloropropene	5.678	75	255714	93.746	ug/l	99
37) Ethyl Acetate	4.702	43	300700	103.517	ug/l	99
38) Carbon Tetrachloride	5.666	117	287048	93.555	ug/l	99
39) Methylcyclohexane	7.367	83	359758	99.352	ug/l	98
40) Benzene	6.025	78	785626	92.946	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
 Data File : VX044652.D
 Acq On : 15 Jan 2025 13:14
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: Jan 16 01:01:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 00:57:36 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
 Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	168481	100.920	ug/l	98
42) 1,2-Dichloroethane	6.074	62	292036	94.663	ug/l	99
43) Isopropyl Acetate	6.330	43	512598	95.925	ug/l	99
44) Trichloroethene	7.117	130	199570	98.212	ug/l	96
45) 1,2-Dichloropropane	7.421	63	200877	94.413	ug/l	99
46) Dibromomethane	7.568	93	150112	92.680	ug/l	98
47) Bromodichloromethane	7.812	83	307990	92.723	ug/l	99
48) Methyl methacrylate	7.684	41	248667	100.250	ug/l	100
49) 1,4-Dioxane	7.665	88	101384	1959.056	ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	1372888	464.897	ug/l	100
52) Toluene	8.708	92	474392	91.697	ug/l	100
53) t-1,3-Dichloropropene	8.970	75	330982	96.446	ug/l	98
54) cis-1,3-Dichloropropene	8.360	75	349516	93.927	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	192326	91.885	ug/l	98
56) Ethyl methacrylate	9.110	69	337626	93.812	ug/l	99
57) 1,3-Dichloropropane	9.305	76	325142	91.220	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	817528	462.693	ug/l	99
59) 2-Hexanone	9.427	43	1019940	469.454	ug/l	99
60) Dibromochloromethane	9.519	129	227418	92.570	ug/l	99
61) 1,2-Dibromoethane	9.604	107	202247	92.200	ug/l	100
64) Tetrachloroethene	9.269	164	157748	98.176	ug/l	98
65) Chlorobenzene	10.073	112	510916	94.003	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	184816	93.565	ug/l	100
67) Ethyl Benzene	10.189	91	909208	95.872	ug/l	99
68) m/p-Xylenes	10.299	106	661488	187.963	ug/l	100
69) o-Xylene	10.640	106	329336	92.536	ug/l	99
70) Styrene	10.653	104	547016	95.197	ug/l	100
71) Bromoform	10.799	173	155379	96.648	ug/l #	99
73) Isopropylbenzene	10.957	105	833874	92.779	ug/l	100
74) N-amyl acetate	10.842	43	414046	93.463	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	278679	89.776	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	250918m	96.568	ug/l	
77) Bromobenzene	11.195	156	192700	91.729	ug/l	99
78) n-propylbenzene	11.299	91	956202	96.470	ug/l	100
79) 2-Chlorotoluene	11.360	91	584682	91.317	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	672586	91.926	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	114832	98.354	ug/l	99
82) 4-Chlorotoluene	11.451	91	655250	93.421	ug/l	100
83) tert-Butylbenzene	11.713	119	690320	90.233	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	681398	91.646	ug/l	100
85) sec-Butylbenzene	11.890	105	848327	95.235	ug/l	100
86) p-Isopropyltoluene	12.006	119	695433	94.999	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	353809	97.115	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	347517	96.070	ug/l	99
89) n-Butylbenzene	12.329	91	633246	100.715	ug/l	99
90) Hexachloroethane	12.536	117	149253	90.037	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	340683	91.056	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	70862	93.344	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	233417	104.359	ug/l	99
94) Hexachlorobutadiene	13.725	225	88214	100.406	ug/l	100
95) Naphthalene	13.774	128	849403	98.464	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	232152	100.133	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044652.D
Acq On : 15 Jan 2025 13:14
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Quant Time: Jan 16 01:01:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044652.D
Acq On : 15 Jan 2025 13:14
Operator : JC/MD
Sample : VSTDIC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Quant Time: Jan 16 01:01:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

