

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011619\
 Data File : VX007075.D
 Acq On : 16 Jan 2019 14:47
 Operator : JC/SP
 Sample : K1010-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-01-011519-B

Quant Time: Jan 17 00:22:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	563143	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	877465	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	840533	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	406597	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	307045	52.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.34%	
35) Dibromofluoromethane	5.50	113	269112	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
50) Toluene-d8	8.71	98	1074714	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
62) 4-Bromofluorobenzene	11.13	95	382475	53.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.68%	
Target Compounds						
16) Acetone	2.45	43	24814	8.529	ug/l	98
20) Methylene Chloride	2.86	84	439596	79.774	ug/l	97
43) Isopropyl Acetate	6.46	43	23670	1.835	ug/l	95
95) Naphthalene	13.83	128	34002	1.161	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011619\
Data File : VX007075.D
Acq On : 16 Jan 2019 14:47
Operator : JC/SP
Sample : K1010-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OK-01-011519-B

Quant Time: Jan 17 00:22:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

