

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011620\
 Data File : VX014634.D
 Acq On : 16 Jan 2020 13:37
 Operator : JC/SP
 Sample : L1135-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TE-MW-A-1

Quant Time: Jan 17 06:20:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 15 12:41:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	70692	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	382996	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	342076	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.05	65	156209	31.44	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.80%
60) 4-Bromofluorobenzene	11.13	95	144082	26.32	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	87.73%
63) Toluene-d8	8.71	98	457700	29.62	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.73%

Target Compounds

					Ovalue
37) Trichloroethene	7.22	130	6914	1.455 ug/l	83
61) Tetrachloroethene	9.33	164	14845	2.936 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011620\
Data File : VX014634.D
Acq On : 16 Jan 2020 13:37
Operator : JC/SP
Sample : L1135-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TE-MW-A-1

Quant Time: Jan 17 06:20:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jan 15 12:41:12 2020
Response via : Initial Calibration

