


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX011620\
 Data File : VX014627.D
 Acq On : 16 Jan 2020 09:38
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X011520W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Wed Jan 15 12:41:12 2020



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1674

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	15208	PASS
75	95	30	60	50.5	39241	PASS
95	95	100	100	100.0	77733	PASS
96	95	5	9	6.9	5383	PASS
173	174	0.00	2	1.2	814	PASS
174	95	50	100	85.9	66741	PASS
175	174	5	9	7.7	5160	PASS
176	174	95	101	96.3	64282	PASS
177	176	5	9	7.0	4475	PASS