

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011623\
 Data File : VX033744.D
 Acq On : 16 Jan 2023 12:25
 Operator : JC/MD
 Sample : VX0116MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0116MBL02

Quant Time: Jan 17 00:40:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	199988	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224270	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95235	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86711	60.779	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 121.560%#	
35) Dibromofluoromethane	5.385	113	67478	54.338	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.680%	
50) Toluene-d8	8.647	98	289992	64.092	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 128.180%#	
62) 4-Bromofluorobenzene	11.079	95	89883	53.505	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.020%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011623\
Data File : VX033744.D
Acq On : 16 Jan 2023 12:25
Operator : JC/MD
Sample : VX0116MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0116MBL02

Quant Time: Jan 17 00:40:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

