

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011623\
 Data File : VX033750.D
 Acq On : 16 Jan 2023 16:00
 Operator : JC/MD
 Sample : 01100-05ME
 Misc : 7.01g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WASTEME

Quant Time: Jan 17 00:44:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	174712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	229869	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95290	51.801	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.600%
35) Dibromofluoromethane	5.385	113	78232	54.809	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.620%
50) Toluene-d8	8.647	98	303343	58.328	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	116.660%#
62) 4-Bromofluorobenzene	11.085	95	98040	50.774	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.540%
Target Compounds						
95) Naphthalene	13.780	128	6126	1.022	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011623\
Data File : VX033750.D
Acq On : 16 Jan 2023 16:00
Operator : JC/MD
Sample : 01100-05ME
Misc : 7.01g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WASTEME

Quant Time: Jan 17 00:44:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

