

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011624\
 Data File : VX039767.D
 Acq On : 16 Jan 2024 12:26
 Operator : JC/MD
 Sample : VX0116WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0116WBS01

Quant Time: Jan 17 01:45:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	214268	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	191140	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88687	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	90236	44.249	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.500%
35) Dibromofluoromethane	5.379	113	68511	47.472	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.940%
50) Toluene-d8	8.647	98	250503	46.022	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.040%
62) 4-Bromofluorobenzene	11.079	95	94761	44.520	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.040%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	25039	15.897	ug/l	99
3) Chloromethane	1.295	50	26460	14.710	ug/l	97
4) Vinyl Chloride	1.374	62	25696	15.137	ug/l	95
5) Bromomethane	1.605	94	17241	17.800	ug/l	99
6) Chloroethane	1.685	64	16007	16.587	ug/l	98
7) Trichlorofluoromethane	1.886	101	40109	16.379	ug/l	100
8) Diethyl Ether	2.130	74	14989	16.563	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.331	101	26102	17.193	ug/l	97
10) Methyl Iodide	2.453	142	27364	17.085	ug/l	91
11) Tert butyl alcohol	2.947	59	41966	79.056	ug/l #	94
12) 1,1-Dichloroethene	2.319	96	23486	16.029	ug/l	97
13) Acrolein	2.233	56	36819	103.841	ug/l	99
14) Allyl chloride	2.666	41	45429	15.753	ug/l	98
15) Acrylonitrile	3.056	53	88681	80.875	ug/l	99
16) Acetone	2.374	43	91023	79.367	ug/l	97
17) Carbon Disulfide	2.514	76	60237	14.304	ug/l	97
18) Methyl Acetate	2.703	43	40122	16.722	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	85411	16.420	ug/l	94
20) Methylene Chloride	2.788	84	28494	16.425	ug/l	94
21) trans-1,2-Dichloroethene	3.093	96	25027	16.231	ug/l	93
22) Diisopropyl ether	3.757	45	92492	16.807	ug/l	98
23) Vinyl Acetate	3.721	43	421749	82.445	ug/l	100
24) 1,1-Dichloroethane	3.611	63	49923	16.490	ug/l	99
25) 2-Butanone	4.550	43	134700	81.730	ug/l	99
26) 2,2-Dichloropropane	4.471	77	39240	16.149	ug/l	97
27) cis-1,2-Dichloroethene	4.483	96	29648	16.765	ug/l	97
28) Bromochloromethane	4.898	49	25298	18.434	ug/l	95
29) Tetrahydrofuran	4.995	42	84625	79.774	ug/l	99
30) Chloroform	5.087	83	50024	16.738	ug/l	99
31) Cyclohexane	5.471	56	42567	15.816	ug/l	93
32) 1,1,1-Trichloroethane	5.379	97	41886	16.370	ug/l	99
36) 1,1-Dichloropropene	5.690	75	36347	16.180	ug/l	99
37) Ethyl Acetate	4.715	43	47045	15.698	ug/l	99
38) Carbon Tetrachloride	5.672	117	33055	16.066	ug/l	94
39) Methylcyclohexane	7.379	83	43982	16.305	ug/l	98
40) Benzene	6.031	78	106652	16.767	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011624\
 Data File : VX039767.D
 Acq On : 16 Jan 2024 12:26
 Operator : JC/MD
 Sample : VX0116WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0116WBS01

Quant Time: Jan 17 01:45:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	25597	17.467	ug/l	95
42) 1,2-Dichloroethane	6.086	62	41700	17.243	ug/l	99
43) Isopropyl Acetate	6.336	43	73736	16.773	ug/l	99
44) Trichloroethene	7.123	130	25952	16.647	ug/l	93
45) 1,2-Dichloropropane	7.428	63	28925	16.951	ug/l	100
46) Dibromomethane	7.580	93	20102	17.056	ug/l	97
47) Bromodichloromethane	7.818	83	37284	17.081	ug/l	95
48) Methyl methacrylate	7.690	41	34843	16.124	ug/l	93
49) 1,4-Dioxane	7.653	88	13019	311.365	ug/l #	100
51) 4-Methyl-2-Pentanone	8.568	43	255305	84.322	ug/l	99
52) Toluene	8.714	92	65826	17.326	ug/l	95
53) t-1,3-Dichloropropene	8.976	75	39314	16.973	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	43375	17.031	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	27643	17.089	ug/l	97
56) Ethyl methacrylate	9.116	69	44849	17.131	ug/l	98
57) 1,3-Dichloropropane	9.305	76	47516	17.334	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	115318	93.188	ug/l	95
59) 2-Hexanone	9.427	43	204790	85.461	ug/l	99
60) Dibromochloromethane	9.519	129	26444	17.335	ug/l	100
61) 1,2-Dibromoethane	9.610	107	29372	17.534	ug/l	99
64) Tetrachloroethene	9.275	164	21660	17.426	ug/l	99
65) Chlorobenzene	10.080	112	70133	17.617	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	24082	17.765	ug/l	98
67) Ethyl Benzene	10.189	91	128726	17.417	ug/l	97
68) m/p-Xylenes	10.299	106	95107	34.989	ug/l	97
69) o-Xylene	10.640	106	46170	17.152	ug/l	96
70) Styrene	10.653	104	76726	17.449	ug/l	99
71) Bromoform	10.799	173	16699	16.406	ug/l #	98
73) Isopropylbenzene	10.964	105	123983	17.559	ug/l	99
74) N-amyl acetate	10.842	43	61668	16.755	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	47814	17.733	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	49531	21.542	ug/l	92
77) Bromobenzene	11.195	156	28128	17.891	ug/l	98
78) n-propylbenzene	11.305	91	148098	17.329	ug/l	98
79) 2-Chlorotoluene	11.360	91	88642	16.782	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	104752	17.468	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	11349	16.480	ug/l	95
82) 4-Chlorotoluene	11.451	91	102732	17.147	ug/l	99
83) tert-Butylbenzene	11.713	119	101080	17.655	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	104406	17.451	ug/l	99
85) sec-Butylbenzene	11.890	105	129691	17.450	ug/l	99
86) p-Isopropyltoluene	12.006	119	103697	17.457	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	52826	17.529	ug/l	100
88) 1,4-Dichlorobenzene	12.043	146	53489	17.185	ug/l	99
89) n-Butylbenzene	12.329	91	95554	16.655	ug/l	97
90) Hexachloroethane	12.536	117	16361	16.327	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	52357	17.573	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	10830	16.392	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	31186	17.128	ug/l	97
94) Hexachlorobutadiene	13.725	225	13513	17.854	ug/l	98
95) Naphthalene	13.774	128	109063	16.272	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	31515	17.150	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011624\
Data File : VX039767.D
Acq On : 16 Jan 2024 12:26
Operator : JC/MD
Sample : VX0116WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01

Quant Time: Jan 17 01:45:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion:168 Resp: 121736
Ion Ratio Lower Upper
168 100
99 61.5 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 15.897 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 85 Resp: 25039
Ion Ratio Lower Upper
85 100
87 34.2 16.7 50.1









#3
Chloromethane
Concen: 14.710 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 50 Resp: 26460
Ion Ratio Lower Upper
50 100
52 33.3 25.5 38.3



#4
Vinyl Chloride
Concen: 15.137 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 62 Resp: 25696
Ion Ratio Lower Upper
62 100
64 34.7 25.4 38.0









#5
Bromomethane
Concen: 17.800 ug/l
RT: 1.605 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 94 Resp: 17241
Ion Ratio Lower Upper
94 100
96 92.2 74.6 111.8



#6
Chloroethane
Concen: 16.587 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 64 Resp: 16007
Ion Ratio Lower Upper
64 100
66 32.3 25.1 37.7









#7
 Trichlorofluoromethane
 Concen: 16.379 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0116WBS01


Tgt Ion:101 Resp: 40109
 Ion Ratio Lower Upper
 101 100
 103 64.0 51.4 77.2



#8
 Diethyl Ether
 Concen: 16.563 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26

Tgt Ion: 74 Resp: 14989
 Ion Ratio Lower Upper
 74 100
 45 101.6 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 17.193 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion:101 Resp: 26102
Ion Ratio Lower Upper
101 100
85 44.4 37.0 55.6
151 70.5 58.6 87.8



#10
Methyl Iodide
Concen: 17.085 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion:142 Resp: 27364
Ion Ratio Lower Upper
142 100
127 44.6 41.6 62.4
141 12.8 11.6 17.4









#11
Tert butyl alcohol
Concen: 79.056 ug/l
RT: 2.947 min Scan# 307
Delta R.T. -0.012 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 59 Resp: 41966
Ion Ratio Lower Upper
59 100
57 9.5 6.0 9.0#



#12
1,1-Dichloroethene
Concen: 16.029 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 96 Resp: 23486
Ion Ratio Lower Upper
96 100
61 176.6 144.6 216.8
98 65.4 50.4 75.6





#13

Acrolein

Concen: 103.841 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0116WBS01

Tgt Ion: 56 Resp: 36819

Ion Ratio Lower Upper

56 100

55 71.8 57.1 85.7



#14

Allyl chloride

Concen: 15.753 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26




Tgt Ion: 41 Resp: 45429


Ion Ratio Lower Upper

41 100

39 67.8 55.8 83.6

76 30.6 24.3 36.5









#15
Acrylonitrile
Concen: 80.875 ug/l
RT: 3.056 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 53 Resp: 88681
Ion Ratio Lower Upper
53 100
52 82.7 66.0 99.0
51 36.3 29.8 44.6



#16
Acetone
Concen: 79.367 ug/l
RT: 2.374 min Scan# 211
Delta R.T. -0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 43 Resp: 91023
Ion Ratio Lower Upper
43 100
58 29.0 21.9 32.9





#17

Carbon Disulfide

Concen: 14.304 ug/l

RT: 2.514 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0116WBS01

Tgt Ion: 76 Resp: 60237

Ion Ratio Lower Upper

76 100

78 9.9 7.0 10.6



#18

Methyl Acetate

Concen: 16.722 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion: 43 Resp: 40122

Ion Ratio Lower Upper

43 100

74 22.9 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 16.420 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 73 Resp: 85411
Ion Ratio Lower Upper
73 100
57 24.3 17.3 25.9



#20
Methylene Chloride
Concen: 16.425 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 84 Resp: 28494
Ion Ratio Lower Upper
84 100
49 138.2 103.6 155.4
51 41.9 33.8 50.6
86 65.7 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 16.231 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 96 Resp: 25027
Ion Ratio Lower Upper
96 100
61 137.1 119.2 178.8
98 60.9 48.7 73.1



#22
Diisopropyl ether
Concen: 16.807 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 45 Resp: 92492
Ion Ratio Lower Upper
45 100
43 89.2 69.7 104.5
87 25.6 20.8 31.2
59 10.9 9.0 13.4





#23

Vinyl Acetate

Concen: 82.445 ug/l

RT: 3.721 min Scan# 41

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0116WBS01

Tgt Ion: 43 Resp: 421749

Ion Ratio Lower Upper

43 100

86 9.7 7.8 11.8



#24

1,1-Dichloroethane

Concen: 16.490 ug/l

RT: 3.611 min Scan# 414

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26




Tgt Ion: 63 Resp: 49923


Ion Ratio Lower Upper

63 100

98 6.0 3.3 9.8

100 4.0 2.1 6.5









#25
2-Butanone
Concen: 81.730 ug/l
RT: 4.550 min Scan# 568
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 43 Resp: 134700
Ion Ratio Lower Upper
43 100
72 25.0 19.4 29.2



#26
2,2-Dichloropropane
Concen: 16.149 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 77 Resp: 39240
Ion Ratio Lower Upper
77 100
97 22.9 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 16.765 ug/l
 RT: 4.483 min Scan# 511
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0116WBS01


Tgt Ion: 96	Resp: 29648
Ion Ratio	Lower Upper
96 100	
61 151.8	0.0 312.2
98 62.8	0.0 129.0



#28
 Bromochloromethane
 Concen: 18.434 ug/l
 RT: 4.898 min Scan# 625
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26

Tgt Ion: 49	Resp: 25298
Ion Ratio	Lower Upper
49 100	
129 1.4	0.0 3.8
130 67.6	57.5 86.3





#29

Tetrahydrofuran

Concen: 79.774 ug/l

RT: 4.995 min Scan# 64

Delta R.T. -0.006 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion: 42 Resp: 84625

Ion Ratio Lower Upper

42 100

72 42.4 34.2 51.4

71 39.2 32.2 48.2



#30

Chloroform

Concen: 16.738 ug/l

RT: 5.087 min Scan# 656

Delta R.T. -0.006 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion: 83 Resp: 50024

Ion Ratio Lower Upper

83 100

85 63.4 50.3 75.5









#31
Cyclohexane
Concen: 15.816 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion:	56	Resp:	42567
Ion	Ratio	Lower	Upper
56	100		
69	28.5	23.4	35.2
84	74.2	65.9	98.9



#32
1,1,1-Trichloroethane
Concen: 16.370 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion:	97	Resp:	41886
Ion	Ratio	Lower	Upper
97	100		
99	63.6	51.1	76.7
61	47.1	36.4	54.6





#33

1,2-Dichloroethane-d4

Concen: 44.249 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0116WBS01

Tgt Ion: 65 Resp: 90236

Ion Ratio Lower Upper

65 100

67 52.6 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26




Tgt Ion: 114 Resp: 214268


Ion Ratio Lower Upper

114 100

63 21.6 0.0 45.0

88 16.5 0.0 36.6









#35
Dibromofluoromethane
Concen: 47.472 ug/l
RT: 5.379 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion:113 Resp: 68511
Ion Ratio Lower Upper
113 100
111 103.8 82.9 124.3
192 16.4 13.3 19.9



#36
1,1-Dichloropropene
Concen: 16.180 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 75 Resp: 36347
Ion Ratio Lower Upper
75 100
110 34.6 17.0 50.9
77 31.4 24.7 37.1





#37

Ethyl Acetate

Concen: 15.698 ug/l

RT: 4.715 min Scan# 594

Delta R.T. 0.006 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26





Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01

Tgt Ion: 43	Resp: 47045
Ion Ratio	Lower Upper
43 100	
61 13.8	10.6 16.0
70 10.5	8.3 12.5



#38

Carbon Tetrachloride

Concen: 16.066 ug/l




RT: 5.672 min Scan# 752


Delta R.T. -0.006 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Tgt Ion: 117	Resp: 33055
Ion Ratio	Lower Upper
117 100	
119 91.6	78.2 117.2
121 27.4	24.4 36.6





#39

Methylcyclohexane

Concen: 16.305 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion: 83 Resp: 43982

Ion Ratio Lower Upper

83 100

55 82.5 64.6 97.0

98 45.1 36.7 55.1



#40

Benzene

Concen: 16.767 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.006 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion: 78 Resp: 106652

Ion Ratio Lower Upper

78 100


77 23.0 19.3 28.9








#41
Methacrylonitrile
Concen: 17.467 ug/l
RT: 4.910 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01







Tgt Ion:	41	Resp:	25597
Ion	Ratio	Lower	Upper
41	100		
39	55.0	45.2	67.8
67	71.0	52.6	78.8
52	31.0	22.7	34.1



#42
1,2-Dichloroethane
Concen: 17.243 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion:	62	Resp:	41700
Ion	Ratio	Lower	Upper
62	100		
98	8.3	0.0	16.0









#43
Isopropyl Acetate
Concen: 16.773 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 43 Resp: 73736
Ion Ratio Lower Upper
43 100
61 18.4 15.2 22.8
87 11.8 9.8 14.6



#44
Trichloroethene
Concen: 16.647 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion:130 Resp: 25952
Ion Ratio Lower Upper
130 100
95 109.6 0.0 204.4









#45
1,2-Dichloropropane
Concen: 16.951 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 63 Resp: 28925
Ion Ratio Lower Upper
63 100
65 30.7 24.3 36.5



#46
Dibromomethane
Concen: 17.056 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 93 Resp: 20102
Ion Ratio Lower Upper
93 100
95 82.2 66.6 100.0
174 83.2 69.4 104.0









#47
Bromodichloromethane
Concen: 17.081 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 83	Resp:	37284	
Ion Ratio	Lower	Upper	
83	100		
85	66.4	49.8	74.6
127	8.2	7.1	10.7



#48
Methyl methacrylate
Concen: 16.124 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 41	Resp:	34843	
Ion	Ratio	Lower	Upper
41	100		
69	80.5	59.2	88.8
39	63.7	47.7	71.5





#49

1,4-Dioxane

Concen: 311.365 ug/l

RT: 7.653 min Scan# 1077

Delta R.T. -0.006 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion: 88 Resp: 13019

Ion Ratio Lower Upper

88 100

43 37.8 0.0 0.0#

58 75.2 60.1 90.1



#50

Toluene-d8

Concen: 46.022 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion: 98 Resp: 250503

Ion Ratio Lower Upper

98 100

100 64.3 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 84.322 ug/l
RT: 8.568 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 43 Resp: 255305
Ion Ratio Lower Upper
43 100
58 37.2 30.4 45.6



#52
Toluene
Concen: 17.326 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 92 Resp: 65826
Ion Ratio Lower Upper
92 100
91 173.8 134.2 201.2









#53
 t-1,3-Dichloropropene
 Concen: 16.973 ug/l
 RT: 8.976 min Scan# 1194
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0116WBS01


Tgt Ion: 75 Resp: 39314
 Ion Ratio Lower Upper
 75 100
 77 31.1 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 17.031 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26

Tgt Ion: 75 Resp: 43375
 Ion Ratio Lower Upper
 75 100
 77 32.5 25.3 37.9
 39 56.0 46.7 70.1









#55
1,1,2-Trichloroethane
Concen: 17.089 ug/l
RT: 9.147 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 97	Resp: 27643
Ion Ratio	Lower Upper
97 100	
83 83.1	65.8 98.8
85 56.8	41.6 62.4
99 61.1	47.8 71.8



#56
Ethyl methacrylate
Concen: 17.131 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 69	Resp: 44849
Ion Ratio	Lower Upper
69 100	
41 71.6	56.8 85.2
39 42.5	32.0 48.0





#57

1,3-Dichloropropane

Concen: 17.334 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0116WBS01

Tgt Ion: 76 Resp: 47516

Ion Ratio Lower Upper

76 100

78 32.1 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 93.188 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion: 63 Resp: 115318

Ion Ratio Lower Upper

63 100

106 25.6 22.5 33.7









#59
2-Hexanone
Concen: 85.461 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 43 Resp: 204790
Ion Ratio Lower Upper
43 100
58 51.8 26.2 78.5



#60
Dibromochloromethane
Concen: 17.335 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 129 Resp: 26444
Ion Ratio Lower Upper
129 100
127 77.9 38.8 116.4





#61

1,2-Dibromoethane

Concen: 17.534 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0116WBS01

Tgt Ion:107 Resp: 29372

Ion Ratio Lower Upper

107 100

109 95.0 75.2 112.8



#62

4-Bromofluorobenzene

Concen: 44.520 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26




Tgt Ion: 95 Resp: 94761


Ion Ratio Lower Upper

95 100

174 68.1 0.0 135.6

176 65.4 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion:117 Resp: 191140
Ion Ratio Lower Upper
117 100
82 61.1 46.2 69.4
119 32.0 25.5 38.3



#64
Tetrachloroethene
Concen: 17.426 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion:164 Resp: 21660
Ion Ratio Lower Upper
164 100
166 128.6 102.7 154.1
129 103.2 80.4 120.6
131 95.9 76.5 114.7





#65

Chlorobenzene

Concen: 17.617 ug/l

RT: 10.080 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0116WBS01

Tgt Ion:112 Resp: 70133

Ion Ratio Lower Upper

112 100

114 31.9 25.7 38.5



#66

1,1,1,2-Tetrachloroethane

Concen: 17.765 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26




Tgt Ion:131 Resp: 24082


Ion Ratio Lower Upper

131 100

133 95.6 47.1 141.4

119 67.5 32.2 96.6









#67
Ethyl Benzene
Concen: 17.417 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.006 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 91 Resp: 128726
Ion Ratio Lower Upper
91 100
106 29.2 24.9 37.3



#68
m/p-Xylenes
Concen: 34.989 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion:106 Resp: 95107
Ion Ratio Lower Upper
106 100
91 212.1 165.5 248.3









#69
o-Xylene
Concen: 17.152 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion:106 Resp: 46170
Ion Ratio Lower Upper
106 100
91 224.8 108.9 326.8



#70
Styrene
Concen: 17.449 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion:104 Resp: 76726
Ion Ratio Lower Upper
104 100
78 54.4 43.6 65.4
103 55.2 43.6 65.4





#71

Bromoform

Concen: 16.406 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion:173 Resp: 16699

Ion Ratio Lower Upper

173 100

175 49.7 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26




Tgt Ion:152 Resp: 88687


Ion Ratio Lower Upper

152 100

115 74.6 43.3 129.9

150 162.9 0.0 345.0





#73

Isopropylbenzene

Concen: 17.559 ug/l

RT: 10.964 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0116WBS01

Tgt Ion:105 Resp: 123983

Ion Ratio Lower Upper

105 100

120 26.3 12.9 38.6



#74

N-ethyl acetate

Concen: 16.755 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Tgt Ion: 43 Resp: 61668




Ion Ratio Lower Upper


43 100

70 38.5 32.9 49.3

55 23.5 19.3 28.9

61 22.1 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 17.733 ug/l

RT: 11.213 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion: 83 Resp: 47814

Ion Ratio Lower Upper

83 100

131 9.1 5.3 15.8

85 63.2 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 21.542 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion: 75 Resp: 49531

Ion Ratio Lower Upper

75 100

77 33.7 19.1 57.5





#77

Bromobenzene

Concen: 17.891 ug/l

RT: 11.195 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion:156 Resp: 28128

Ion Ratio Lower Upper

156 100

77 180.1 88.4 265.3

158 93.9 48.0 144.0



#78

n-propylbenzene

Concen: 17.329 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion: 91 Resp: 148098

Ion Ratio Lower Upper

91 100

120 22.0 11.5 34.4





#79

2-Chlorotoluene

Concen: 16.782 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :





VX0116WBS01

Tgt Ion: 91 Resp: 88642

Ion Ratio Lower Upper

91 100

126 33.1 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 17.468 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion: 105 Resp: 104752

Ion Ratio Lower Upper

105 100

120 46.1 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 16.480 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26




Instrument :
MSVOA_X
ClientSampleId :
VX0116WBS01


Tgt Ion: 75 Resp: 11349
Ion Ratio Lower Upper
75 100
53 120.6 90.5 135.7
89 47.3 36.9 55.3



#82
4-Chlorotoluene
Concen: 17.147 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX039767.D
Acq: 16 Jan 2024 12:26

Tgt Ion: 91 Resp: 102732
Ion Ratio Lower Upper
91 100
126 28.8 14.2 42.6





#83

tert-Butylbenzene

Concen: 17.655 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion:119 Resp: 101080

Ion Ratio Lower Upper

119 100

91 63.3 31.2 93.6

134 21.9 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 17.451 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion:105 Resp: 104406

Ion Ratio Lower Upper

105 100

120 43.3 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 17.450 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0116WBS01


Tgt Ion:105 Resp: 129691
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 17.457 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26

Tgt Ion:119 Resp: 103697
 Ion Ratio Lower Upper
 119 100
 134 26.8 12.8 38.3
 91 27.5 13.1 39.3









#87
 1,3-Dichlorobenzene
 Concen: 17.529 ug/l
 RT: 11.969 min Scan# 1785
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0116WBS01


Tgt Ion:146 Resp: 52826
 Ion Ratio Lower Upper
 146 100
 111 44.4 22.1 66.1
 148 63.8 31.8 95.3



#88
 1,4-Dichlorobenzene
 Concen: 17.185 ug/l
 RT: 12.043 min Scan# 1797
 Delta R.T. 0.000 min
 Lab File: VX039767.D
 Acq: 16 Jan 2024 12:26

Tgt Ion:146 Resp: 53489
 Ion Ratio Lower Upper
 146 100
 111 44.0 21.3 63.7
 148 63.2 31.8 95.4





#89

n-Butylbenzene

Concen: 16.655 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion: 91 Resp: 95554

Ion Ratio Lower Upper

91 100

92 52.5 27.4 82.0

134 24.3 12.6 37.8



#90

Hexachloroethane

Concen: 16.327 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX039767.D




Acq: 16 Jan 2024 12:26


Tgt Ion: 117 Resp: 16361

Ion Ratio Lower Upper

117 100

201 64.7 32.9 98.6





#91

1,2-Dichlorobenzene

Concen: 17.573 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion:146 Resp: 52357

Ion Ratio Lower Upper

146 100

111 45.3 22.4 67.0

148 64.3 31.6 95.0



#92

1,2-Dibromo-3-Chloropropane

Concen: 16.392 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. -0.006 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26




Tgt Ion: 75 Resp: 10830


Ion Ratio Lower Upper

75 100

155 70.7 36.1 108.5

157 89.9 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 17.128 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion:180 Resp: 31186

Ion Ratio Lower Upper

180 100

182 95.0 48.3 144.8

145 32.3 18.0 54.0



#94

Hexachlorobutadiene

Concen: 17.854 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26




Tgt Ion:225 Resp: 13513


Ion Ratio Lower Upper

225 100

223 65.1 31.1 93.2

227 65.0 32.5 97.5





#95

Naphthalene

Concen: 16.272 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Instrument :

MSVOA_X

ClientSampleId :

VX0116WBS01





Tgt Ion:128 Resp: 109063

Ion Ratio Lower Upper

128 100

127 12.6 10.3 15.5

129 10.7 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 17.150 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.006 min

Lab File: VX039767.D

Acq: 16 Jan 2024 12:26

Tgt Ion:180 Resp: 31515

Ion Ratio Lower Upper

180 100

182 92.0 46.9 140.6

145 34.9 17.8 53.5

