

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011624\
 Data File : VX039769.D
 Acq On : 16 Jan 2024 13:16
 Operator : JC/MD
 Sample : PB158447TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158447TB

Quant Time: Jan 17 01:47:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95680	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	182164	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	85035	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73472	45.840	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.680%
35) Dibromofluoromethane	5.385	113	58113	47.364	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.720%
50) Toluene-d8	8.647	98	227525	49.168	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.340%
62) 4-Bromofluorobenzene	11.079	95	92238	50.972	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.940%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	28187	31.271	ug/l	92
20) Methylene Chloride	2.788	84	5175	3.796	ug/l	93
37) Ethyl Acetate	4.733	43	2763	1.084	ug/l #	86
43) Isopropyl Acetate	6.336	43	102405	27.399	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011624\
Data File : VX039769.D
Acq On : 16 Jan 2024 13:16
Operator : JC/MD
Sample : PB158447TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158447TB

Quant Time: Jan 17 01:47:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

