

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011624\
 Data File : VX039774.D
 Acq On : 16 Jan 2024 15:09
 Operator : JC/MD
 Sample : P1139-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-314-16-18

Quant Time: Jan 17 01:50:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	94476	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	174823	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	86523	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73536	46.465	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.920%
35) Dibromofluoromethane	5.385	113	57369	48.721	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	8.647	98	220523	49.656	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.320%
62) 4-Bromofluorobenzene	11.079	95	92360	53.182	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.360%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	32342	36.337	ug/l	92
20) Methylene Chloride	2.794	84	4972	3.693	ug/l	94
43) Isopropyl Acetate	6.336	43	89289	24.893	ug/l	99
68) m/p-Xylenes	10.299	106	5338	2.110	ug/l	99
69) o-Xylene	10.640	106	2910	1.162	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	9171	1.571	ug/l	99
95) Naphthalene	13.774	128	26459	4.046	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011624\
Data File : VX039774.D
Acq On : 16 Jan 2024 15:09
Operator : JC/MD
Sample : P1139-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-314-16-18

Quant Time: Jan 17 01:50:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

