

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011624\
 Data File : VX039776.D
 Acq On : 16 Jan 2024 15:54
 Operator : JC/MD
 Sample : P1149-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-GW-58-60

Quant Time: Jan 17 01:50:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	96722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	179825	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	88325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73596	45.423	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.840%
35) Dibromofluoromethane	5.385	113	58170	48.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.060%
50) Toluene-d8	8.647	98	227417	49.784	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.560%
62) 4-Bromofluorobenzene	11.079	95	93560	52.375	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.740%
Target Compounds						
16) Acetone	2.374	43	3985	4.373	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011624\
Data File : VX039776.D
Acq On : 16 Jan 2024 15:54
Operator : JC/MD
Sample : P1149-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-GW-58-60

Quant Time: Jan 17 01:50:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

