

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011625\
 Data File : VX044671.D
 Acq On : 16 Jan 2025 14:18
 Operator :
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 17 01:31:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011625W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jan 17 01:21:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	30872	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	158542	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.049	117	144585	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.946	65	89155	29.914	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.700%
60) 4-Bromofluorobenzene	11.079	95	81087	28.874	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	96.233%
63) Toluene-d8	8.647	98	238445	30.321	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.067%
Target Compounds						Qvalue
62) Toluene	8.720	91	3176	0.334	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011625\
Data File : VX044671.D
Acq On : 16 Jan 2025 14:18
Operator :
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 17 01:31:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011625W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jan 17 01:21:41 2025
Response via : Initial Calibration

