

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX011719\
 Data File : VX007115.D
 Acq On : 18 Jan 2019 03:49
 Operator : JC/SP
 Sample : K1167-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CLOVER-YARD-A

Quant Time: Jan 18 05:09:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	588659	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	906419	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	875139	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	423613	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	309733	50.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.70%	
35) Dibromofluoromethane	5.51	113	279112	48.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.16%	
50) Toluene-d8	8.71	98	1100211	51.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.56%	
62) 4-Bromofluorobenzene	11.13	95	394083	53.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.40%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.33	50	2598	0.462	ug/l #	88
8) Diethyl Ether	2.20	74	1237	0.318	ug/l	83
11) Tert butyl alcohol	3.05	59	538	0.371	ug/l #	1
13) Acrolein	2.29	56	852	4.429	ug/l #	13
14) Allyl chloride	2.85	41	2773279	299.098	ug/l #	25
16) Acetone	2.45	43	28832	9.480	ug/l	94
18) Methyl Acetate	2.79	43	15279	1.735	ug/l #	89
21) trans-1,2-Dichloroethene	3.17	96	1835	0.304	ug/l #	70
25) 2-Butanone	4.70	43	7522	1.729	ug/l #	89
29) Tetrahydrofuran	5.16	42	1859	0.657	ug/l #	59
36) 1,1-Dichloropropene	5.66	75	37638	4.776	ug/l #	48
38) Carbon Tetrachloride	5.67	117	53239	6.715	ug/l #	14
43) Isopropyl Acetate	6.46	43	26931	2.021	ug/l	98
48) Methyl methacrylate	7.69	41	13874	2.027	ug/l #	12
49) 1,4-Dioxane	7.74	88	69	Below Cal	#	1
51) 4-Methyl-2-Pentanone	8.62	43	126852	14.811	ug/l #	48
54) cis-1,3-Dichloropropene	8.62	75	63611	7.183	ug/l #	63
57) 1,3-Dichloropropane	9.54	76	15353	1.515	ug/l #	45
59) 2-Hexanone	9.54	43	197516	30.263	ug/l #	51
68) m/p-Xylenes	10.35	106	4885	0.393	ug/l	96
76) 1,2,3-Trichloropropane	11.13	75	190495	22.191	ug/l #	40
81) trans-1,4-Dichloro-2-buten	11.13	75	190495	65.685	ug/l #	10
84) 1,2,4-Trimethylbenzene	11.81	105	11063	0.441	ug/l	96
88) 1,4-Dichlorobenzene	12.08	146	829	0.204	ug/l #	1
90) Hexachloroethane	12.57	117	213	2.955	ug/l #	6
95) Naphthalene	13.83	128	18792	0.616	ug/l #	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007115.D
Acq On : 18 Jan 2019 03:49
Operator : JC/SP
Sample : K1167-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CLOVER-YARD-A

Quant Time: Jan 18 05:09:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :


CLOVER-YARD-A

Tot Ion:168 Resp: 588659

Ion Ratio Lower Upper

168 100

99 50.1 41.8 62.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

200000



150000

100000

50000

0

Time-->



#3

Chloromethane

Concen: 0.462 ug/l

RT: 1.33 min Scan# 29

Delta R.T. -0.00 min

Lab File: VX007115.D


Acq: 18 Jan 2019 03:49

Tgt Ion: 50 Resp: 2598

Ion Ratio Lower Upper

50 100

52 25.0 25.4 38.2#



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0

1.33

2000


1500


1000

500

0

Time-->









#8
Diethyl Ether
Concen: 0.318 ug/l
RT: 2.20 min Scan# 172
Delta R.T. 0.01 min
Lab File: VX007115.D
Acq: 18 Jan 2019 03:49




Instrument :
MSVOA_X
ClientSampleId :
CLOVER-YARD-A


Tot Ion: 74 Resp: 1237
Ion Ratio Lower Upper
74 100
45 76.4 46.3 138.9



#11
Tert butyl alcohol
Concen: 0.371 ug/l
RT: 3.05 min Scan# 311
Delta R.T. -0.02 min
Lab File: VX007115.D
Acq: 18 Jan 2019 03:49

Tgt Ion: 59 Resp: 538
Ion Ratio Lower Upper
59 100
57 60.2 8.5 12.7#





#13

Acrolein

Concen: 4.429 ug/l

RT: 2.29 min Scan# 187

Delta R.T. -0.01 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :





CLOVER-YARD-A

Tot Ion: 56 Resp: 852

Ion Ratio Lower Upper

56 100

55 0.0 58.2 87.4#



#14

Allyl chloride

Concen: 299.098 ug/l

RT: 2.85 min Scan# 279

Delta R.T. 0.13 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49




Tgt Ion: 41 Resp: 2773279


Ion Ratio Lower Upper

41 100

39 0.8 53.6 80.4#

76 0.0 27.4 41.2#





#16

Acetone

Concen: 9.480 ug/l

RT: 2.45 min Scan# 212

Delta R.T. -0.00 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :





CLOVER-YARD-A

Tot Ion: 43 Resp: 28832

Ion Ratio Lower Upper

43 100

58 34.8 25.0 37.6



#18

Methyl Acetate

Concen: 1.735 ug/l

RT: 2.79 min Scan# 268

Delta R.T. 0.01 min

Lab File: VX007115.D




Acq: 18 Jan 2019 03:49


Tgt Ion: 43 Resp: 15279

Ion Ratio Lower Upper

43 100

74 28.8 18.6 28.0#





#21

trans-1,2-Dichloroethene

Concen: 0.304 ug/l

RT: 3.17 min Scan# 330

Delta R.T. -0.00 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :

CLOVER-YARD-A





Tot Ion: 96 Resp: 1835

Ion Ratio Lower Upper

96 100

61 88.9 103.7 155.5#

98 49.0 51.0 76.6#



#25

2-Butanone

Concen: 1.729 ug/l

RT: 4.70 min Scan# 582

Delta R.T. -0.00 min

Lab File: VX007115.D




Acq: 18 Jan 2019 03:49


Tgt Ion: 43 Resp: 7522

Ion Ratio Lower Upper

43 100

72 22.1 22.4 33.6#





#29

Tetrahydrofuran

Concen: 0.657 ug/l

RT: 5.16 min Scan# 657

Delta R.T. 0.01 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :

CLOVER-YARD-A





Tot Ion: 42 Resp: 1859

Ion Ratio Lower Upper

42 100

72 69.3 38.9 58.3#

71 10.4 36.2 54.4#



#33

1,2-Dichloroethane-d4

Concen: 50.347 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX007115.D




Acq: 18 Jan 2019 03:49


Tgt Ion: 65 Resp: 309733

Ion Ratio Lower Upper

65 100

67 53.8 0.0 105.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :

CLOVER-YARD-A


Tot Ion:114 Resp: 906419

Ion Ratio Lower Upper

114 100

63 18.1 0.0 34.8




88 14.6 0.0 28.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.084 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49


Tgt Ion:113 Resp: 279112

Ion Ratio Lower Upper

113 100

111 101.4 81.6 122.4



192 22.1 16.7 25.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 4.776 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.14 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :

CLOVER-YARD-A


Tot Ion: 75 Resp: 37638

Ion Ratio Lower Upper

75 100

110 10.0 20.2 60.5#




77 0.0 24.8 37.2#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0



#38

Carbon Tetrachloride

Concen: 6.715 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.12 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49


Tgt Ion: 117 Resp: 53239

Ion Ratio Lower Upper

117 100

119 5.0 79.4 119.2#



121 0.0 24.0 36.0#




Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V





#43

Isopropyl Acetate

Concen: 2.021 ug/l

RT: 6.46 min Scan# 870

Delta R.T. -0.00 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :

CLOVER-YARD-A


Tot Ion: 43 Resp: 26931

Ion Ratio Lower Upper

43 100

61 21.7 16.8 25.2




87 16.5 12.4 18.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#48

Methyl methacrylate

Concen: 2.027 ug/l

RT: 7.69 min Scan# 1072

Delta R.T. -0.09 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49


Tgt Ion: 41 Resp: 13874

Ion Ratio Lower Upper

41 100

69 1.0 72.4 108.6#



39 0.0 45.8 68.8#




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: Below Cal

RT: 7.74 min Scan# 1080

Delta R.T. -0.01 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :

CLOVER-YARD-A





Tot Ion: 88 Resp: 69

Ion Ratio Lower Upper

88 100

43 0.0 22.2 33.4#

58 1389.9 51.0 76.4#



#50

Toluene-d8

Concen: 51.782 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX007115.D




Acq: 18 Jan 2019 03:49


Tgt Ion: 98 Resp: 1100211

Ion Ratio Lower Upper

98 100

100 64.2 52.0 78.0





#51

4-Methyl-2-Pentanone

Concen: 14.811 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. -0.03 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :





CLOVER-YARD-A

Tot Ion: 43 Resp: 126852

Ion Ratio Lower Upper

43 100

58 7.8 31.8 47.6#



#54

cis-1,3-Dichloropropene

Concen: 7.183 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. 0.18 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49




Tgt Ion: 75 Resp: 63611


Ion Ratio Lower Upper

75 100

77 0.6 26.2 39.2#

39 30.7 36.7 55.1#





#57

1,3-Dichloropropane

Concen: 1.515 ug/l

RT: 9.54 min Scan# 1376

Delta R.T. 0.17 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :





CLOVER-YARD-A

Tot Ion: 76 Resp: 15353

Ion Ratio Lower Upper

76 100

78 1.4 26.0 39.0#



#59

2-Hexanone

Concen: 30.263 ug/l

RT: 9.54 min Scan# 1376

Delta R.T. 0.05 min

Lab File: VX007115.D




Acq: 18 Jan 2019 03:49


Tgt Ion: 43 Resp: 197516

Ion Ratio Lower Upper

43 100

58 19.9 27.7 83.1#





#62

4-Bromofluorobenzene

Concen: 53.203 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :

CLOVER-YARD-A


Tot Ion: 95 Resp: 394083

Ion Ratio Lower Upper

95 100

174 92.4 0.0 182.2

176 90.9 0.0 178.8




Abundance Ion 94.90 (94.60 to 95.60): VX006951.D (-1631) (-)

Ion 173.80 (173.50 to 174.50): VX006951.D (-1631) (-)

Ion 175.80 (175.50 to 176.50): VX006951.D (-1631) (-)

Time-->




Abundance

Ion 94.90 (94.60 to 95.60): VX007115.D (-1588) (-)

Ion 173.80 (173.50 to 174.50): VX007115.D (-1588) (-)

Ion 175.80 (175.50 to 176.50): VX007115.D (-1588) (-)

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.12 min Scan# 1470

Delta R.T. -0.00 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49


Tgt Ion: 117 Resp: 875139

Ion Ratio Lower Upper

117 100

82 50.3 41.0 61.6

119 32.8 25.4 38.0




Abundance Ion 116.90 (116.60 to 117.60): VX006951.D (-1462) (-)

Ion 82.00 (81.70 to 82.70): VX006951.D (-1462) (-)

Ion 118.90 (118.60 to 119.60): VX006951.D (-1462) (-)

Time-->




Abundance

Ion 116.90 (116.60 to 117.60): VX007115.D (-1421) (-)

Ion 82.00 (81.70 to 82.70): VX007115.D (-1421) (-)

Ion 118.90 (118.60 to 119.60): VX007115.D (-1421) (-)

Time-->



#68

m/p-Xvlenes

Concen: 0.393 ug/l

RT: 10.35 min Scan# 1509

Delta R.T. -0.01 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :


CLOVER-YARD-A

Tot Ion:106 Resp: 4885

Ion Ratio Lower Upper




106 100

91 189.2 155.8 233.6



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VX0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49


Tgt Ion:152 Resp: 423613

Ion Ratio Lower Upper

152 100

115 54.5 39.1 117.2



150 157.3 0.0 344.8




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#76

1,2,3-Trichloropropane

Concen: 22.191 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.16 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :





CLOVER-YARD-A

Tot Ion: 75 Resp: 190495

Ion Ratio Lower Upper

75 100

77 1.1 18.5 55.5#



#81

trans-1,4-Dichloro-2-butene

Concen: 65.685 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.06 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49




Tgt Ion: 75 Resp: 190495


Ion Ratio Lower Upper

75 100

53 0.2 79.7 119.5#

89 0.0 38.0 57.0#





#84

1,2,4-Trimethylbenzene

Concen: 0.441 ug/l

RT: 11.81 min Scan# 1748

Delta R.T. 0.01 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :


CLOVER-YARD-A

Tot Ion:105 Resp: 11063

Ion Ratio Lower Upper




105 100

120 43.6 23.2 69.6



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#88

1,4-Dichlorobenzene

Concen: 0.204 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.02 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49


Tgt Ion:146 Resp: 829

Ion Ratio Lower Upper

146 100

111 639.3 18.1 54.1#



148 211.5 31.9 95.9#




Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V





#90

Hexachloroethane

Concen: 2.955 ug/l

RT: 12.57 min Scan# 1872

Delta R.T. -0.03 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49

Instrument :

MSVOA_X

ClientSampleId :


CLOVER-YARD-A

Tot Ion:117 Resp: 213

Ion Ratio Lower Upper




117 100

201 0.0 42.9 128.7#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#95

Naphthalene

Concen: 0.616 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. -0.00 min

Lab File: VX007115.D

Acq: 18 Jan 2019 03:49


Tgt Ion:128 Resp: 18792

Ion Ratio Lower Upper

128 100

127 14.0 10.1 15.1

129 15.2 8.6 13.0#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

