

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX011719\
 Data File : VX007082.D
 Acq On : 17 Jan 2019 12:16
 Operator : JC/SP
 Sample : K1010-05ME
 Misc : 4.67g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-01-011519-BME

Quant Time: Jan 18 09:26:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	509405	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	767749	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	723887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	409063	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	267620	50.27	ug/l	0.00
Spiked Amount			Recovery	=	100.54%	
35) Dibromofluoromethane	5.51	113	224327	45.63	ug/l	0.00
Spiked Amount			Recovery	=	91.26%	
50) Toluene-d8	8.71	98	910250	50.58	ug/l	0.00
Spiked Amount			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	11.14	95	351376	56.01	ug/l	0.00
Spiked Amount			Recovery	=	112.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
67) Ethyl Benzene	10.25	91	42295	1.623	ug/l	98
68) m/p-Xylenes	10.36	106	107550	10.473	ug/l	100
69) o-Xylene	10.70	106	73479	7.337	ug/l	99
73) Isopropylbenzene	11.02	105	91565	3.174	ug/l	94
78) n-propylbenzene	11.36	91	330819	10.447	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	509938	21.158	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	1992336	82.204	ug/l	99
85) sec-Butylbenzene	11.94	105	327813	11.415	ug/l	80
86) p-Isopropyltoluene	12.07	119	311081	12.110	ug/l	99
89) n-Butylbenzene	12.39	91	702671	32.088	ug/l #	60
95) Naphthalene	13.83	128	883850	29.989	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007082.D
Acq On : 17 Jan 2019 12:16
Operator : JC/SP
Sample : K1010-05ME
Misc : 4.67g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OK-01-011519-BME

Quant Time: Jan 18 09:26:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

