

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
 Data File : VX007093.D
 Acq On : 17 Jan 2019 17:09
 Operator : JC/SP
 Sample : K1172-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST010MW005-190114

Quant Time: Jan 18 09:49:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	598791	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	913756	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	853582	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	421694	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	312825	49.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.98%	
35) Dibromofluoromethane	5.51	113	279520	47.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.54%	
50) Toluene-d8	8.71	98	1088100	50.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.60%	
62) 4-Bromofluorobenzene	11.13	95	388250	51.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.98%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007093.D
Acq On : 17 Jan 2019 17:09
Operator : JC/SP
Sample : K1172-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST010MW005-190114

Quant Time: Jan 18 09:49:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

