

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
 Data File : VX007107.D
 Acq On : 18 Jan 2019 00:15
 Operator : JC/SP
 Sample : K1153-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR200044-RBR200027

Quant Time: Jan 18 07:11:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	590997	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	900331	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	850207	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	405232	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	308279	49.91	ug/l	0.00
Spiked Amount			Recovery	=	99.82%	
35) Dibromofluoromethane	5.50	113	276437	47.95	ug/l	0.00
Spiked Amount			Recovery	=	95.90%	
50) Toluene-d8	8.71	98	1091744	51.73	ug/l	0.00
Spiked Amount			Recovery	=	103.46%	
62) 4-Bromofluorobenzene	11.13	95	379044	51.52	ug/l	0.00
Spiked Amount			Recovery	=	103.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	20666	6.768	ug/l	95
20) Methylene Chloride	2.86	84	761858	132.111	ug/l	96
25) 2-Butanone	4.70	43	10742	2.459	ug/l #	85
43) Isopropyl Acetate	6.46	43	34169	2.581	ug/l	99
95) Naphthalene	13.83	128	49441	1.693	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007107.D
Acq On : 18 Jan 2019 00:15
Operator : JC/SP
Sample : K1153-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR200044-RBR200027

Quant Time: Jan 18 07:11:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

