

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
 Data File : VX007108.D
 Acq On : 18 Jan 2019 00:42
 Operator : JC/SP
 Sample : K1156-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-5

Quant Time: Jan 18 07:13:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	577147	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	894899	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	841530	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	400008	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	308889	51.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.42%	
35) Dibromofluoromethane	5.51	113	276893	48.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.64%	
50) Toluene-d8	8.71	98	1075355	51.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.52%	
62) 4-Bromofluorobenzene	11.13	95	372617	50.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.90%	
Target Compounds						
16) Acetone	2.45	43	26829	8.997	ug/l	96
18) Methyl Acetate	2.79	43	13721	1.589	ug/l	97
20) Methylene Chloride	2.86	84	421708	74.634	ug/l	96
25) 2-Butanone	4.70	43	12275	2.877	ug/l	93
43) Isopropyl Acetate	6.46	43	34209	2.600	ug/l	99
95) Naphthalene	13.83	128	41266	1.432	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007108.D
Acq On : 18 Jan 2019 00:42
Operator : JC/SP
Sample : K1156-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-5

Quant Time: Jan 18 07:13:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

