

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
 Data File : VX007115.D
 Acq On : 18 Jan 2019 03:49
 Operator : JC/SP
 Sample : K1167-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CLOVER-YARD-A

Manual Integrations
 APPROVED

MMDadoda
 1/18/2019 10:22:43 AM

Quant Time: Jan 18 07:28:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	588659	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	906419	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	875139	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	423613	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	309733	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
35) Dibromofluoromethane	5.51	113	279112	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
50) Toluene-d8	8.71	98	1100211	51.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.56%	
62) 4-Bromofluorobenzene	11.13	95	394083	53.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.40%	
Target Compounds						
16) Acetone	2.45	43	28832	9.480	ug/l	94
18) Methyl Acetate	2.79	43	15279	1.735	ug/l #	89
20) Methylene Chloride	2.84	84	27685098m	4840.165	ug/l	
25) 2-Butanone	4.70	43	7522	1.729	ug/l #	89
43) Isopropyl Acetate	6.46	43	26931	2.021	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007115.D
Acq On : 18 Jan 2019 03:49
Operator : JC/SP
Sample : K1167-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CLOVER-YARD-A

Manual Integrations
APPROVED

MMDadoda
1/18/2019 10:22:43 AM

Quant Time: Jan 18 07:28:08 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

