

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
 Data File : VX007116.D
 Acq On : 18 Jan 2019 04:15
 Operator : JC/SP
 Sample : K1167-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CLOVER-YARD-B

Manual Integrations
 APPROVED

MMDadoda
 1/18/2019 10:22:37 AM

Quant Time: Jan 18 07:30:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	590051	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	922736	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	879396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	422968	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	316533	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
35) Dibromofluoromethane	5.51	113	280744	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	8.71	98	1119508	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
62) 4-Bromofluorobenzene	11.13	95	396502	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
Target Compounds						
16) Acetone	2.45	43	33032	10.835	ug/l	94
18) Methyl Acetate	2.80	43	18417	2.086	ug/l #	80
20) Methylene Chloride	2.84	84	32038739m	5588.184	ug/l	
25) 2-Butanone	4.71	43	6230	1.428	ug/l #	85
43) Isopropyl Acetate	6.46	43	24515	1.807	ug/l	96
95) Naphthalene	13.83	128	42748	1.403	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007116.D
Acq On : 18 Jan 2019 04:15
Operator : JC/SP
Sample : K1167-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CLOVER-YARD-B

Manual Integrations
APPROVED

MMDadoda
1/18/2019 10:22:37 AM

Quant Time: Jan 18 07:30:10 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

