

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
 Data File : VX007120.D
 Acq On : 18 Jan 2019 06:02
 Operator : JC/SP
 Sample : K1157-26
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-06-011519

Quant Time: Jan 18 07:41:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	549940	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	856546	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	813007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	387717	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	300153	52.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.44%	
35) Dibromofluoromethane	5.50	113	263955	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
50) Toluene-d8	8.71	98	1045842	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
62) 4-Bromofluorobenzene	11.13	95	362711	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
Target Compounds						
16) Acetone	2.45	43	41108	14.468	ug/l	98
18) Methyl Acetate	2.78	43	20295	2.467	ug/l	99
20) Methylene Chloride	2.86	84	39223	6.769	ug/l	93
43) Isopropyl Acetate	6.46	43	18832	1.495	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007120.D
Acq On : 18 Jan 2019 06:02
Operator : JC/SP
Sample : K1157-26
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-06-011519

Quant Time: Jan 18 07:41:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

