

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
 Data File : VX007121.D
 Acq On : 18 Jan 2019 06:28
 Operator : JC/SP
 Sample : K1157-28
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-07-011519

Quant Time: Jan 18 07:43:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	569332	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	883468	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	827925	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	397585	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	301999	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
35) Dibromofluoromethane	5.50	113	269446	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
50) Toluene-d8	8.71	98	1076534	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
62) 4-Bromofluorobenzene	11.13	95	375183	51.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.94%	
Target Compounds						
16) Acetone	2.45	43	81467	27.696	ug/l	97
18) Methyl Acetate	2.78	43	21830	2.563	ug/l	93
20) Methylene Chloride	2.86	84	30812	4.998	ug/l	92
43) Isopropyl Acetate	6.46	43	13983	1.077	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007121.D
Acq On : 18 Jan 2019 06:28
Operator : JC/SP
Sample : K1157-28
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-07-011519

Quant Time: Jan 18 07:43:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

