

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
 Data File : VX007103.D
 Acq On : 17 Jan 2019 22:02
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 18 07:03:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.411	0.407	1.0	82	0.00
3 P	Chloromethane	0.478	0.446	6.7	82	0.00
4 C	Vinyl Chloride	0.514	0.482	6.2#	83	0.00
5 T	Bromomethane	0.549	0.479	12.8	84	0.00
6 T	Chloroethane	0.359	0.327	8.9	89	0.00
7 T	Trichlorofluoromethane	0.831	0.795	4.3	88	0.00
8 T	Diethyl Ether	0.331	0.319	3.6	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.497	0.469	5.6	91	0.00
10 T	Methyl Iodide	0.671	0.693	-3.3	86	0.00
11 T	Tert butyl alcohol	0.123	0.127	-3.3	98	0.01
12 CM	1,1-Dichloroethene	0.460	0.432	6.1#	87	0.00
13 T	Acrolein	0.056	0.053	5.4	88	0.00
14 T	Allyl chloride	0.788	0.762	3.3	88	0.00
15 T	Acrylonitrile	0.283	0.293	-3.5	97	0.00
16 T	Acetone	0.258	0.252	2.3	93	0.00
17 T	Carbon Disulfide	1.183	0.982	17.0	76	0.00
18 T	Methyl Acetate	0.748	0.791	-5.7	102	0.00
19 T	Methyl tert-butyl Ether	1.595	1.594	0.1	92	0.00
20 T	Methylene Chloride	0.548	0.508	7.3	92	0.00
21 T	trans-1,2-Dichloroethene	0.512	0.471	8.0	88	0.00
22 T	Diisopropyl ether	1.481	1.465	1.1	91	0.00
23 T	Vinyl Acetate	1.249	1.211	3.0	89	0.00
24 P	1,1-Dichloroethane	0.828	0.803	3.0	89	0.00
25 T	2-Butanone	0.370	0.376	-1.6	93	0.00
26 T	2,2-Dichloropropane	0.637	0.543	14.8	79	0.00
27 T	cis-1,2-Dichloroethene	0.565	0.537	5.0	88	0.00
28 T	Bromochloromethane	0.363	0.352	3.0	83	0.00
29 T	Tetrahydrofuran	0.240	0.245	-2.1	94	0.00
30 C	Chloroform	0.867	0.847	2.3#	91	0.00
31 T	Cyclohexane	0.737	0.676	8.3	84	0.00
32 T	1,1,1-Trichloroethane	0.735	0.716	2.6	88	0.00
33 S	1,2-Dichloroethane-d4	0.523	0.515	1.5	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.320	0.315	1.6	83	0.00
36 T	1,1-Dichloropropene	0.435	0.406	6.7	87	0.00
37 T	Ethyl Acetate	0.466	0.468	-0.4	92	0.00
38 T	Carbon Tetrachloride	0.437	0.418	4.3	88	0.00
39 T	Methylcyclohexane	0.554	0.496	10.5	84	0.00
40 TM	Benzene	1.326	1.254	5.4	88	0.00
41 T	Methacrylonitrile	0.268	0.258	3.7	94	0.00
42 TM	1,2-Dichloroethane	0.458	0.437	4.6	90	0.00
43 T	Isopropyl Acetate	0.735	0.752	-2.3	92	0.00
44 TM	Trichloroethene	0.397	0.365	8.1	87	0.00
45 C	1,2-Dichloropropane	0.333	0.328	1.5#	91	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX011719\
 Data File : VX007103.D
 Acq On : 17 Jan 2019 22:02
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 18 07:03:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.234	0.228	2.6	90	0.00
47 T	Bromodichloromethane	0.399	0.402	-0.8	90	0.00
48 T	Methyl methacrylate	0.378	0.382	-1.1	93	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	94	0.00
50 S	Toluene-d8	1.172	1.166	0.5	83	0.00
51 T	4-Methyl-2-Pentanone	0.472	0.499	-5.7	96	0.00
52 CM	Toluene	0.853	0.815	4.5#	89	0.00
53 T	t-1,3-Dichloropropene	0.442	0.440	0.5	86	0.00
54 T	cis-1,3-Dichloropropene	0.489	0.492	-0.6	89	0.00
55 T	1,1,2-Trichloroethane	0.341	0.353	-3.5	93	0.00
56 T	Ethyl methacrylate	0.498	0.525	-5.4	95	0.00
57 T	1,3-Dichloropropane	0.559	0.561	-0.4	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.256	0.266	-3.9	88	0.00
59 T	2-Hexanone	0.360	0.378	-5.0	96	0.00
60 T	Dibromochloromethane	0.331	0.357	-7.9	91	0.00
61 T	1,2-Dibromoethane	0.376	0.383	-1.9	92	0.00
62 S	4-Bromofluorobenzene	0.409	0.411	-0.5	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.433	0.400	7.6	91	0.00
65 PM	Chlorobenzene	1.091	1.047	4.0	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.367	0.383	-4.4	93	0.00
67 C	Ethyl Benzene	1.800	1.738	3.4#	92	0.00
68 T	m/p-Xylenes	0.709	0.684	3.5	91	0.00
69 T	o-Xylene	0.692	0.673	2.7	92	0.00
70 T	Styrene	1.096	1.090	0.5	93	0.00
71 P	Bromoform	0.281	0.296	-5.3	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.526	3.470	1.6	93	0.00
74 T	N-amyl acetate	1.420	1.436	-1.1	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.119	1.101	1.6	95	0.00
76 T	1,2,3-Trichloropropane	1.013	1.029	-1.6	95	0.00
77 T	Bromobenzene	0.971	0.933	3.9	92	0.00
78 T	n-propylbenzene	3.870	3.840	0.8	93	0.00
79 T	2-Chlorotoluene	2.330	2.263	2.9	94	0.00
80 T	1,3,5-Trimethylbenzene	2.946	2.831	3.9	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.285	0.300	-5.3	89	0.00
82 T	4-Chlorotoluene	2.698	2.606	3.4	92	0.00
83 T	tert-Butylbenzene	2.939	2.915	0.8	93	0.00
84 T	1,2,4-Trimethylbenzene	2.962	2.913	1.7	92	0.00
85 T	sec-Butylbenzene	3.510	3.456	1.5	93	0.00
86 T	p-Isopropyltoluene	3.140	3.084	1.8	93	0.00
87 T	1,3-Dichlorobenzene	1.756	1.665	5.2	93	0.00
88 T	1,4-Dichlorobenzene	1.806	1.662	8.0	94	0.00
89 T	n-Butylbenzene	2.677	2.633	1.6	92	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
Data File : VX007103.D
Acq On : 17 Jan 2019 22:02
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 18 07:03:10 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.443	0.482	-8.8	94	0.00
91 T	1,2-Dichlorobenzene	1.733	1.649	4.8	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.221	0.242	-9.5	98	0.00
93 T	1,2,4-Trichlorobenzene	1.189	1.170	1.6	94	0.00
94 T	Hexachlorobutadiene	0.528	0.541	-2.5	98	0.00
95 T	Naphthalene	3.602	3.645	-1.2	97	0.00
96 T	1,2,3-Trichlorobenzene	1.181	1.159	1.9	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6