


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011719\
 Data File : VX007102.D
 Acq On : 17 Jan 2019 21:36
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Title : SW846 8260
 Last Update : Tue Jan 08 14:06:59 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1628

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	14200	PASS
75	95	30	60	50.1	41416	PASS
95	95	100	100	100.0	82738	PASS
96	95	5	9	6.6	5497	PASS
173	174	0.00	2	1.2	911	PASS
174	95	50	100	89.7	74256	PASS
175	174	5	9	7.5	5562	PASS
176	174	95	101	96.1	71389	PASS
177	176	5	9	6.9	4935	PASS