

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011720\
 Data File : VX014646.D
 Acq On : 17 Jan 2020 11:36
 Operator : JC/SP
 Sample : VX0117WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0117WBL02

Quant Time: Jan 18 05:32:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	420958	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	649061	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	587576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	268205	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	240459	51.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.34%	
35) Dibromofluoromethane	5.49	113	197588	50.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.58%	
50) Toluene-d8	8.71	98	770732	50.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.10%	
62) 4-Bromofluorobenzene	11.13	95	262118	46.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.66%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011720\
Data File : VX014646.D
Acq On : 17 Jan 2020 11:36
Operator : JC/SP
Sample : VX0117WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0117WBL02

Quant Time: Jan 18 05:32:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

