

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011720\
 Data File : VX014655.D
 Acq On : 17 Jan 2020 15:21
 Operator : JC/SP
 Sample : L1162-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-5-(SETTLED)

Quant Time: Jan 18 05:52:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	429820	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	671771	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	602380	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	271640	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	248134	51.71	ug/l	0.00
Spiked Amount			Recovery	=	103.42%	
35) Dibromofluoromethane	5.49	113	202876	49.89	ug/l	0.00
Spiked Amount			Recovery	=	99.78%	
50) Toluene-d8	8.71	98	800211	50.71	ug/l	0.00
Spiked Amount			Recovery	=	101.42%	
62) 4-Bromofluorobenzene	11.13	95	267446	46.16	ug/l	0.00
Spiked Amount			Recovery	=	92.32%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011720\
Data File : VX014655.D
Acq On : 17 Jan 2020 15:21
Operator : JC/SP
Sample : L1162-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-5-(SETTLED)

Quant Time: Jan 18 05:52:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

