

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX011720\
 Data File : VX014662.D
 Acq On : 17 Jan 2020 18:04
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 18 06:08:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.557	0.596	-7.0	73	0.00
3 P	Chloromethane	0.637	0.640	-0.5	78	0.00
4 C	Vinyl Chloride	0.710	0.733	-3.2#	76	0.00
5 T	Bromomethane	0.471	0.459	2.5	83	0.00
6 T	Chloroethane	0.420	0.402	4.3	78	0.00
7 T	Trichlorofluoromethane	0.806	0.805	0.1	78	0.00
8 T	Diethyl Ether	0.368	0.361	1.9	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.490	0.483	1.4	79	0.00
10 T	Methyl Iodide	0.587	0.655	-11.6	79	0.00
11 T	Tert butyl alcohol	0.117	0.129	-10.3	85	0.00
12 CM	1,1-Dichloroethene	0.494	0.477	3.4#	77	0.00
13 T	Acrolein	0.093	0.100	-7.5	84	0.00
14 T	Allyl chloride	0.867	0.838	3.3	76	0.00
15 T	Acrylonitrile	0.281	0.311	-10.7	87	0.00
16 T	Acetone	0.303	0.286	5.6	78	0.00
17 T	Carbon Disulfide	1.378	1.221	11.4	72	0.00
18 T	Methyl Acetate	0.802	0.939	-17.1	88	0.00
19 T	Methyl tert-butyl Ether	1.585	1.651	-4.2	81	0.00
20 T	Methylene Chloride	0.562	0.589	-4.8	86	0.00
21 T	trans-1,2-Dichloroethene	0.539	0.514	4.6	78	0.00
22 T	Diisopropyl ether	1.675	1.827	-9.1	85	0.00
23 T	Vinyl Acetate	1.297	1.437	-10.8	87	0.00
24 P	1,1-Dichloroethane	0.911	0.965	-5.9	82	0.00
25 T	2-Butanone	0.417	0.442	-6.0	83	0.00
26 T	2,2-Dichloropropane	0.772	0.756	2.1	77	0.00
27 T	cis-1,2-Dichloroethene	0.603	0.607	-0.7	81	0.00
28 T	Bromochloromethane	0.384	0.389	-1.3	84	0.00
29 T	Tetrahydrofuran	0.247	0.280	-13.4	86	0.00
30 C	Chloroform	0.891	0.950	-6.6#	84	0.00
31 T	Cyclohexane	0.853	0.875	-2.6	77	0.00
32 T	1,1,1-Trichloroethane	0.762	0.793	-4.1	81	0.00
33 S	1,2-Dichloroethane-d4	0.558	0.588	-5.4	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.303	0.317	-4.6	85	0.00
36 T	1,1-Dichloropropene	0.470	0.456	3.0	78	0.00
37 T	Ethyl Acetate	0.511	0.539	-5.5	84	0.00
38 T	Carbon Tetrachloride	0.433	0.437	-0.9	79	0.00
39 T	Methylcyclohexane	0.591	0.571	3.4	76	0.00
40 TM	Benzene	1.396	1.429	-2.4	82	0.00
41 T	Methacrylonitrile	0.272	0.299	-9.9	86	0.00
42 TM	1,2-Dichloroethane	0.488	0.494	-1.2	83	0.00
43 T	Isopropyl Acetate	0.825	0.878	-6.4	83	0.00
44 TM	Trichloroethene	0.401	0.392	2.2	79	0.00
45 C	1,2-Dichloropropane	0.356	0.373	-4.8#	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX011720\
 Data File : VX014662.D
 Acq On : 17 Jan 2020 18:04
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 18 06:08:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.248	0.243	2.0	82	0.00
47 T	Bromodichloromethane	0.452	0.478	-5.8	83	0.00
48 T	Methyl methacrylate	0.395	0.436	-10.4	84	0.00
49 T	1,4-Dioxane	0.008	0.009	-12.5	88	0.00
50 S	Toluene-d8	1.175	1.213	-3.2	84	0.00
51 T	4-Methyl-2-Pentanone	0.503	0.564	-12.1	87	0.00
52 CM	Toluene	0.893	0.895	-0.2#	80	0.00
53 T	t-1,3-Dichloropropene	0.526	0.520	1.1	78	0.00
54 T	cis-1,3-Dichloropropene	0.568	0.582	-2.5	79	0.00
55 T	1,1,2-Trichloroethane	0.359	0.371	-3.3	83	0.00
56 T	Ethyl methacrylate	0.550	0.582	-5.8	80	0.00
57 T	1,3-Dichloropropane	0.600	0.623	-3.8	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.187	0.203	-8.6	84	0.00
59 T	2-Hexanone	0.398	0.431	-8.3	83	0.00
60 T	Dibromochloromethane	0.361	0.385	-6.6	81	0.00
61 T	1,2-Dibromoethane	0.380	0.382	-0.5	82	0.00
62 S	4-Bromofluorobenzene	0.431	0.434	-0.7	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.501	0.428	14.6	68	0.00
65 PM	Chlorobenzene	1.085	1.051	3.1	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.389	-0.5	80	0.00
67 C	Ethyl Benzene	1.849	1.876	-1.5#	81	0.00
68 T	m/p-Xylenes	0.714	0.710	0.6	80	0.00
69 T	o-Xylene	0.681	0.685	-0.6	80	0.00
70 T	Styrene	1.175	1.183	-0.7	80	0.00
71 P	Bromoform	0.317	0.320	-0.9	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.511	3.601	-2.6	80	0.00
74 T	N-amyl acetate	1.549	1.645	-6.2	82	0.00
75 P	1,1,2,2-Tetrachloroethane	1.126	1.206	-7.1	87	0.00
76 T	1,2,3-Trichloropropane	1.035	1.013	2.1	85	0.00
77 T	Bromobenzene	0.974	0.931	4.4	79	0.00
78 T	n-propylbenzene	4.021	4.105	-2.1	80	0.00
79 T	2-Chlorotoluene	2.397	2.418	-0.9	81	0.00
80 T	1,3,5-Trimethylbenzene	2.915	3.010	-3.3	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.391	0.396	-1.3	76	0.00
82 T	4-Chlorotoluene	2.813	2.790	0.8	81	0.00
83 T	tert-Butylbenzene	2.754	2.835	-2.9	82	0.00
84 T	1,2,4-Trimethylbenzene	2.946	3.031	-2.9	81	0.00
85 T	sec-Butylbenzene	3.424	3.447	-0.7	79	0.00
86 T	p-Isopropyltoluene	3.177	3.217	-1.3	79	0.00
87 T	1,3-Dichlorobenzene	1.747	1.665	4.7	79	0.00
88 T	1,4-Dichlorobenzene	1.792	1.638	8.6	77	0.00
89 T	n-Butylbenzene	2.905	2.805	3.4	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX011720\
Data File : VX014662.D
Acq On : 17 Jan 2020 18:04
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 18 06:08:39 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.558	0.571	-2.3	80	0.00
91 T	1,2-Dichlorobenzene	1.712	1.664	2.8	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.253	0.263	-4.0	81	0.00
93 T	1,2,4-Trichlorobenzene	1.310	1.133	13.5	74	0.00
94 T	Hexachlorobutadiene	0.610	0.548	10.2	76	0.00
95 T	Naphthalene	3.730	3.475	6.8	75	0.00
96 T	1,2,3-Trichlorobenzene	1.278	1.144	10.5	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6