

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011722\
 Data File : VX026481.D
 Acq On : 17 Jan 2022 11:04
 Operator : JC/MD
 Sample : VX0117MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0117MBL01

Quant Time: Jan 18 00:40:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135268	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92638	49.025	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.060%
35) Dibromofluoromethane	5.391	113	72090	49.234	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.460%
50) Toluene-d8	8.647	98	260752	49.418	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.840%
62) 4-Bromofluorobenzene	11.079	95	96440	51.451	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.900%
Target Compounds						
16) Acetone	2.392	43	2635	2.734	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011722\
Data File : VX026481.D
Acq On : 17 Jan 2022 11:04
Operator : JC/MD
Sample : VX0117MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0117MBL01

Quant Time: Jan 18 00:40:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

