

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011722\
 Data File : VX026482.D
 Acq On : 17 Jan 2022 11:27
 Operator : JC/MD
 Sample : VX0117WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0117WBL01

Quant Time: Jan 18 00:40:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	225333	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	92382	48.911	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.820%
35) Dibromofluoromethane	5.391	113	71949	49.445	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.900%
50) Toluene-d8	8.653	98	260929	49.761	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.520%
62) 4-Bromofluorobenzene	11.079	95	98834	53.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.120%
Target Compounds						
16) Acetone	2.392	43	3438	3.569	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011722\
Data File : VX026482.D
Acq On : 17 Jan 2022 11:27
Operator : JC/MD
Sample : VX0117WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0117WBL01

Quant Time: Jan 18 00:40:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

