

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX011722\
 Data File : VX026492.D
 Acq On : 17 Jan 2022 16:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 18 00:41:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	50.000	43.748	12.5	83	0.00
3 P	Chloromethane	50.000	45.645	8.7	89	0.00
4 C	Vinyl Chloride	50.000	45.886	8.2#	86	0.00
5 T	Bromomethane	50.000	45.958	8.1	89	0.00
6 T	Chloroethane	50.000	51.119	-2.2	98	0.00
7 T	Trichlorofluoromethane	50.000	44.573	10.9	82	0.00
8 T	Diethyl Ether	50.000	45.306	9.4	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	46.172	7.7	85	0.00
10 T	Methyl Iodide	50.000	53.827	-7.7	107	0.00
11 T	Tert butyl alcohol	250.000	244.009	2.4	88	-0.02
12 CM	1,1-Dichloroethene	50.000	45.496	9.0#	84	0.00
13 T	Acrolein	250.000	253.732	-1.5	102	0.00
14 T	Allyl chloride	50.000	47.459	5.1	86	0.00
15 T	Acrylonitrile	250.000	241.739	3.3	91	0.00
16 T	Acetone	250.000	196.184	21.5	72	0.00
17 T	Carbon Disulfide	50.000	43.202	13.6	83	0.00
18 T	Methyl Acetate	50.000	46.687	6.6	89	0.00
19 T	Methyl tert-butyl Ether	50.000	45.697	8.6	85	0.00
20 T	Methylene Chloride	50.000	44.321	11.4	87	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.260	9.5	84	0.00
22 T	Diisopropyl ether	50.000	46.082	7.8	85	0.00
23 T	Vinyl Acetate	250.000	245.075	2.0	88	0.00
24 P	1,1-Dichloroethane	50.000	46.126	7.7	85	0.00
25 T	2-Butanone	250.000	231.041	7.6	84	0.00
26 T	2,2-Dichloropropane	50.000	48.113	3.8	85	0.00
27 T	cis-1,2-Dichloroethene	50.000	46.245	7.5	84	0.00
28 T	Bromochloromethane	50.000	45.951	8.1	84	0.00
29 T	Tetrahydrofuran	250.000	246.065	1.6	92	0.00
30 C	Chloroform	50.000	46.178	7.6#	86	0.00
31 T	Cyclohexane	50.000	44.572	10.9	84	0.00
32 T	1,1,1-Trichloroethane	50.000	47.292	5.4	85	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.267	9.5	84	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	47.171	5.7	85	0.00
36 T	1,1-Dichloropropene	50.000	46.157	7.7	84	0.00
37 T	Ethyl Acetate	50.000	48.858	2.3	88	0.00
38 T	Carbon Tetrachloride	50.000	47.959	4.1	85	0.00
39 T	Methylcyclohexane	50.000	47.416	5.2	86	0.00
40 TM	Benzene	50.000	46.625	6.8	85	0.00
41 T	Methacrylonitrile	50.000	50.276	-0.6	91	0.00
42 TM	1,2-Dichloroethane	50.000	45.869	8.3	83	0.00
43 T	Isopropyl Acetate	50.000	49.731	0.5	88	0.00
44 TM	Trichloroethene	50.000	47.331	5.3	85	0.00
45 C	1,2-Dichloropropane	50.000	47.549	4.9#	85	0.00
46 T	Dibromomethane	50.000	47.636	4.7	86	0.00
47 T	Bromodichloromethane	50.000	48.819	2.4	86	0.00
48 T	Methyl methacrylate	50.000	50.175	-0.3	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011722\
 Data File : VX026492.D
 Acq On : 17 Jan 2022 16:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 18 00:41:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	843.957	15.6	75	-0.01
50 S	Toluene-d8	50.000	47.832	4.3	87	0.00
51 T	4-Methyl-2-Pentanone	250.000	257.327	-2.9	93	0.00
52 CM	Toluene	50.000	48.010	4.0#	87	0.00
53 T	t-1,3-Dichloropropene	50.000	46.738	6.5	89	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.737	-1.5	87	0.00
55 T	1,1,2-Trichloroethane	50.000	48.478	3.0	88	0.00
56 T	Ethyl methacrylate	50.000	51.505	-3.0	89	0.00
57 T	1,3-Dichloropropane	50.000	49.051	1.9	89	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	248.263	0.7	89	0.00
59 T	2-Hexanone	250.000	253.002	-1.2	89	0.00
60 T	Dibromochloromethane	50.000	51.848	-3.7	90	0.00
61 T	1,2-Dibromoethane	50.000	50.129	-0.3	90	0.00
62 S	4-Bromofluorobenzene	50.000	51.659	-3.3	93	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	45.730	8.5	88	0.00
65 PM	Chlorobenzene	50.000	46.334	7.3	90	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.391	5.2	89	0.00
67 C	Ethyl Benzene	50.000	47.027	5.9#	89	0.00
68 T	m/p-Xylenes	100.000	96.247	3.8	91	0.00
69 T	o-Xylene	50.000	47.939	4.1	90	0.00
70 T	Styrene	50.000	49.605	0.8	91	0.00
71 P	Bromoform	50.000	47.214	5.6	96	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	44.854	10.3	91	0.00
74 T	N-ethyl acetate	50.000	50.646	-1.3	96	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.150	5.7	96	0.00
76 T	1,2,3-Trichloropropane	50.000	48.646	2.7	96	0.00
77 T	Bromobenzene	50.000	45.334	9.3	91	0.00
78 T	n-propylbenzene	50.000	46.345	7.3	93	0.00
79 T	2-Chlorotoluene	50.000	45.123	9.8	92	0.00
80 T	1,3,5-Trimethylbenzene	50.000	45.919	8.2	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.540	2.9	104	0.00
82 T	4-Chlorotoluene	50.000	45.791	8.4	93	0.00
83 T	tert-Butylbenzene	50.000	46.239	7.5	92	0.00
84 T	1,2,4-Trimethylbenzene	50.000	45.815	8.4	91	0.00
85 T	sec-Butylbenzene	50.000	47.253	5.5	93	0.00
86 T	p-Isopropyltoluene	50.000	48.084	3.8	95	0.00
87 T	1,3-Dichlorobenzene	50.000	46.147	7.7	95	0.00
88 T	1,4-Dichlorobenzene	50.000	45.581	8.8	93	0.00
89 T	n-Butylbenzene	50.000	49.254	1.5	96	0.00
90 T	Hexachloroethane	50.000	46.833	6.3	99	0.00
91 T	1,2-Dichlorobenzene	50.000	47.184	5.6	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.516	-3.0	99	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.949	4.1	95	0.00
94 T	Hexachlorobutadiene	50.000	48.388	3.2	98	0.00
95 T	Naphthalene	50.000	49.144	1.7	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011722\
Data File : VX026492.D
Acq On : 17 Jan 2022 16:48
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 18 00:41:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.207	3.6	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6