

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011724\
 Data File : VX039788.D
 Acq On : 17 Jan 2024 09:49
 Operator : JC/MD
 Sample : VX0117MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0117MBL01

Quant Time: Jan 18 02:37:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	97523	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	179791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	173230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	81656	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74496	45.601	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.200%
35) Dibromofluoromethane	5.379	113	58440	48.259	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.520%
50) Toluene-d8	8.647	98	221621	48.524	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.079	95	87851	49.188	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011724\
Data File : VX039788.D
Acq On : 17 Jan 2024 09:49
Operator : JC/MD
Sample : VX0117MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0117MBL01

Quant Time: Jan 18 02:37:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

