

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011724\
 Data File : VX039795.D
 Acq On : 17 Jan 2024 12:30
 Operator : JC/MD
 Sample : P1156-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUMOGW-9-37-40

Quant Time: Jan 18 02:41:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	199126	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	193295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82881	45.954	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.900%
35) Dibromofluoromethane	5.379	113	64749	48.277	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.560%
50) Toluene-d8	8.647	98	245266	48.487	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.980%
62) 4-Bromofluorobenzene	11.079	95	98038	49.562	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.120%
Target Compounds						
					Qvalue	
30) Chloroform	5.092	83	115458	43.680	ug/l	97
52) Toluene	8.720	92	1440	0.408	ug/l	81
64) Tetrachloroethene	9.268	164	727	0.578	ug/l	93
68) m/p-Xylenes	10.305	106	1797	0.654	ug/l	93
69) o-Xylene	10.640	106	585	0.215	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	1656	0.260	ug/l	97
95) Naphthalene	13.780	128	2315	0.324	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011724\
Data File : VX039795.D
Acq On : 17 Jan 2024 12:30
Operator : JC/MD
Sample : P1156-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-9-37-40

Quant Time: Jan 18 02:41:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

