

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011724\
 Data File : VX039797.D
 Acq On : 17 Jan 2024 13:15
 Operator : JC/MD
 Sample : P1156-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUMOGW-8-10-13

Quant Time: Jan 18 02:41:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	88345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	165173	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	165979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83239	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69083	46.681	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.360%
35) Dibromofluoromethane	5.379	113	53566	48.149	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.300%
50) Toluene-d8	8.647	98	207362	49.420	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.840%
62) 4-Bromofluorobenzene	11.079	95	83934	51.154	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.300%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	2165	2.601	ug/l #	88
64) Tetrachloroethene	9.281	164	631	0.585	ug/l	92
68) m/p-Xylenes	10.305	106	554	0.235	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011724\
Data File : VX039797.D
Acq On : 17 Jan 2024 13:15
Operator : JC/MD
Sample : P1156-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-8-10-13

Quant Time: Jan 18 02:41:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

