

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011819\
 Data File : VX007130.D
 Acq On : 18 Jan 2019 13:09
 Operator : JC/SP
 Sample : K1194-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 18 22:18:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	532931	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	833302	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	768337	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	376347	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	287103	51.55	ug/l	0.00
Spiked Amount			Recovery	=	103.10%	
35) Dibromofluoromethane	5.51	113	257889	48.33	ug/l	0.00
Spiked Amount			Recovery	=	96.66%	
50) Toluene-d8	8.71	98	986195	50.49	ug/l	0.00
Spiked Amount			Recovery	=	100.98%	
62) 4-Bromofluorobenzene	11.13	95	356621	52.37	ug/l	0.00
Spiked Amount			Recovery	=	104.74%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011819\
Data File : VX007130.D
Acq On : 18 Jan 2019 13:09
Operator : JC/SP
Sample : K1194-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 18 22:18:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

