

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011819\
 Data File : VX007134.D
 Acq On : 18 Jan 2019 14:55
 Operator : JC/SP
 Sample : K1199-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TB01-20190114

Quant Time: Jan 21 09:54:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	545111	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	844014	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	790520	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	382891	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	293701	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
35) Dibromofluoromethane	5.51	113	269243	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	8.71	98	1009935	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
62) 4-Bromofluorobenzene	11.13	95	356525	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011819\
Data File : VX007134.D
Acq On : 18 Jan 2019 14:55
Operator : JC/SP
Sample : K1199-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TB01-20190114

Quant Time: Jan 21 09:54:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

