

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011819\
 Data File : VX007139.D
 Acq On : 18 Jan 2019 17:08
 Operator : JC/SP
 Sample : K1195-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS038RW016-190116

Quant Time: Jan 18 22:57:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	559526	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	858693	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	805094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	401389	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	292624	50.04	ug/l	0.00
Spiked Amount			Recovery	=	100.08%	
35) Dibromofluoromethane	5.51	113	263804	47.97	ug/l	0.00
Spiked Amount			Recovery	=	95.94%	
50) Toluene-d8	8.71	98	1031283	51.24	ug/l	0.00
Spiked Amount			Recovery	=	102.48%	
62) 4-Bromofluorobenzene	11.13	95	372919	53.14	ug/l	0.00
Spiked Amount			Recovery	=	106.28%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.41	62	32069	5.573	ug/l #	80
16) Acetone	2.45	43	3738	1.293	ug/l #	82
27) cis-1,2-Dichloroethene	4.60	96	28496	4.508	ug/l	88
39) Methylcyclohexane	7.45	83	4287	0.450	ug/l #	75
40) Benzene	6.15	78	12889	0.566	ug/l #	77
65) Chlorobenzene	10.14	112	323397	18.417	ug/l	97
73) Isopropylbenzene	11.02	105	7896	0.279	ug/l	97
83) tert-Butylbenzene	11.77	119	40317	1.709	ug/l	88
85) sec-Butylbenzene	11.94	105	69809	2.477	ug/l #	55
88) 1,4-Dichlorobenzene	12.10	146	7159	0.679	ug/l #	42
91) 1,2-Dichlorobenzene	12.39	146	5380	0.387	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011819\
Data File : VX007139.D
Acq On : 18 Jan 2019 17:08
Operator : JC/SP
Sample : K1195-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS038RW016-190116

Quant Time: Jan 18 22:57:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

