

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011822\
 Data File : VX026497.D
 Acq On : 18 Jan 2022 13:34
 Operator : JC/MD
 Sample : VX0118WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0118WBL02

Quant Time: Jan 18 13:58:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	151650	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	256454	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89518	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	102596	48.430	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.860%
35) Dibromofluoromethane	5.391	113	80961	48.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.780%
50) Toluene-d8	8.653	98	293881	49.244	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.480%
62) 4-Bromofluorobenzene	11.079	95	104084	49.096	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011822\
Data File : VX026497.D
Acq On : 18 Jan 2022 13:34
Operator : JC/MD
Sample : VX0118WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0118WBL02

Quant Time: Jan 18 13:58:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

