

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011822\
 Data File : VX026500.D
 Acq On : 18 Jan 2022 14:53
 Operator : JC/MD
 Sample : N1161-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DSN-001-SEWER-OUTLET-001

Quant Time: Jan 19 07:31:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 10 12:50:55 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/19/2022
 Supervised By : Mahesh Dadoda 01/19/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	18653	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	102901	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	86056	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	44019	29.348	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	97.833%	
60) 4-Bromofluorobenzene	11.079	95	40231	28.012	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	93.367%	
63) Toluene-d8	8.653	98	119187	30.148	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	100.500%	
Target Compounds						
15) Acetone	2.386	58	17210	93.508	ug/l	Qvalue 98
25) Chloroform	5.105	83	3186m	1.489	ug/l	
41) Bromodichloromethane	7.830	83	741	0.454	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011822\
Data File : VX026500.D
Acq On : 18 Jan 2022 14:53
Operator : JC/MD
Sample : N1161-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DSN-001-SEWER-OUTLET-001

Quant Time: Jan 19 07:31:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 10 12:50:55 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/19/2022
Supervised By : Mahesh Dadoda 01/19/2022

