

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011823\
 Data File : VX033784.D
 Acq On : 18 Jan 2023 19:24
 Operator : JC/MD
 Sample : 01209-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP-03

Quant Time: Jan 19 03:34:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 18 14:40:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	15389	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	103892	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	110821	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	38278	32.142	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	107.133%
60) 4-Bromofluorobenzene	11.079	95	57610	29.079	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	96.933%
63) Toluene-d8	8.647	98	106441	30.105	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.333%
Target Compounds						
15) Acetone	2.386	58	1073	4.589	ug/l	Qvalue 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011823\
Data File : VX033784.D
Acq On : 18 Jan 2023 19:24
Operator : JC/MD
Sample : 01209-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP-03

Quant Time: Jan 19 03:34:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jan 18 14:40:07 2023
Response via : Initial Calibration

