

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
 Data File : VX039807.D
 Acq On : 18 Jan 2024 10:49
 Operator : JC/MD
 Sample : VX0118WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0118WBL01

Quant Time: Jan 19 00:36:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	91885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175474	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	171852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	84883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73128	47.510	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.020%
35) Dibromofluoromethane	5.385	113	56883	48.129	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.260%
50) Toluene-d8	8.647	98	217886	48.880	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.760%
62) 4-Bromofluorobenzene	11.079	95	88040	50.507	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
Data File : VX039807.D
Acq On : 18 Jan 2024 10:49
Operator : JC/MD
Sample : VX0118WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0118WBL01

Quant Time: Jan 19 00:36:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

