

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
 Data File : VX039811.D
 Acq On : 18 Jan 2024 12:23
 Operator : JC/MD
 Sample : P1150-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-306-0-2

Quant Time: Jan 19 00:38:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	89843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	171143	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80802	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71492	47.503	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.000%
35) Dibromofluoromethane	5.385	113	54214	47.031	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.060%
50) Toluene-d8	8.647	98	212096	48.785	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	83101	48.880	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.760%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	33247	39.280	ug/l	90
20) Methylene Chloride	2.788	84	6255	4.886	ug/l	95
43) Isopropyl Acetate	6.330	43	68963	19.640	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
Data File : VX039811.D
Acq On : 18 Jan 2024 12:23
Operator : JC/MD
Sample : P1150-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-306-0-2

Quant Time: Jan 19 00:38:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

