

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
 Data File : VX039812.D
 Acq On : 18 Jan 2024 12:45
 Operator : JC/MD
 Sample : P1150-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-306-25-27

Quant Time: Jan 19 00:38:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	97526	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	182491	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	185350	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	92477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76294	46.700	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.400%
35) Dibromofluoromethane	5.385	113	58610	47.683	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.360%
50) Toluene-d8	8.647	98	232371	50.125	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.240%
62) 4-Bromofluorobenzene	11.079	95	95050	52.432	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.860%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	39529	43.023	ug/l	95
20) Methylene Chloride	2.794	84	5208	3.747	ug/l #	89
43) Isopropyl Acetate	6.336	43	68052	18.175	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
Data File : VX039812.D
Acq On : 18 Jan 2024 12:45
Operator : JC/MD
Sample : P1150-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-306-25-27

Quant Time: Jan 19 00:38:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

