

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
 Data File : VX039813.D
 Acq On : 18 Jan 2024 13:08
 Operator : JC/MD
 Sample : P1159-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-011724

Quant Time: Jan 19 00:39:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	93052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	175299	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170674	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	82701	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72254	46.354	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.700%		
35) Dibromofluoromethane	5.385	113	56783	48.092	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.180%		
50) Toluene-d8	8.647	98	217490	48.840	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.680%		
62) 4-Bromofluorobenzene	11.079	95	86871	49.886	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	99.780%		
Target Compounds						
16) Acetone	2.374	43	13518	15.420	ug/l	95
20) Methylene Chloride	2.794	84	3489	2.631	ug/l	95
43) Isopropyl Acetate	6.336	43	74101	20.603	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
Data File : VX039813.D
Acq On : 18 Jan 2024 13:08
Operator : JC/MD
Sample : P1159-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-011724

Quant Time: Jan 19 00:39:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

