

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX011824\
 Data File : VX039823.D
 Acq On : 18 Jan 2024 16:53
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 19 00:44:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.647	0.589	9.0	69	0.00
3 P	Chloromethane	0.739	0.662	10.4	71	0.00
4 C	Vinyl Chloride	0.697	0.668	4.2#	76	0.00
5 T	Bromomethane	0.398	0.423	-6.3	84	0.00
6 T	Chloroethane	0.396	0.411	-3.8	78	0.00
7 T	Trichlorofluoromethane	1.006	1.051	-4.5	84	0.00
8 T	Diethyl Ether	0.372	0.379	-1.9	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.624	0.640	-2.6	81	0.00
10 T	Methyl Iodide	0.658	0.730	-10.9	82	0.00
11 T	Tert butyl alcohol	0.218	0.215	1.4	77	0.00
12 CM	1,1-Dichloroethene	0.602	0.619	-2.8#	81	0.00
13 T	Acrolein	0.147	0.129	12.2	74	0.00
14 T	Allyl chloride	1.184	1.166	1.5	77	0.00
15 T	Acrylonitrile	0.450	0.447	0.7	77	0.00
16 T	Acetone	0.471	0.422	10.4	69	0.00
17 T	Carbon Disulfide	1.730	1.689	2.4	78	0.00
18 T	Methyl Acetate	0.985	1.020	-3.6	81	0.00
19 T	Methyl tert-butyl Ether	2.136	2.173	-1.7	81	0.00
20 T	Methylene Chloride	0.713	0.707	0.8	80	0.00
21 T	trans-1,2-Dichloroethene	0.633	0.651	-2.8	82	0.00
22 T	Diisopropyl ether	2.260	2.303	-1.9	81	0.00
23 T	Vinyl Acetate	2.101	2.191	-4.3	79	0.00
24 P	1,1-Dichloroethane	1.243	1.270	-2.2	80	0.00
25 T	2-Butanone	0.677	0.655	3.2	75	0.00
26 T	2,2-Dichloropropane	0.998	1.012	-1.4	80	0.00
27 T	cis-1,2-Dichloroethene	0.726	0.760	-4.7	81	0.00
28 T	Bromochloromethane	0.564	0.560	0.7	77	0.00
29 T	Tetrahydrofuran	0.436	0.423	3.0	75	0.00
30 C	Chloroform	1.228	1.309	-6.6#	84	0.00
31 T	Cyclohexane	1.105	1.102	0.3	78	0.00
32 T	1,1,1-Trichloroethane	1.051	1.125	-7.0	83	0.00
33 S	1,2-Dichloroethane-d4	0.838	0.772	7.9	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.337	0.333	1.2	78	0.00
36 T	1,1-Dichloropropene	0.524	0.546	-4.2	83	0.00
37 T	Ethyl Acetate	0.699	0.695	0.6	79	0.00
38 T	Carbon Tetrachloride	0.480	0.527	-9.8	83	0.00
39 T	Methylcyclohexane	0.629	0.650	-3.3	80	0.00
40 TM	Benzene	1.484	1.563	-5.3	82	0.00
41 T	Methacrylonitrile	0.342	0.367	-7.3	77	0.00
42 TM	1,2-Dichloroethane	0.564	0.627	-11.2	85	0.00
43 T	Isopropyl Acetate	1.026	1.089	-6.1	79	0.00
44 TM	Trichloroethene	0.364	0.402	-10.4	85	0.00
45 C	1,2-Dichloropropane	0.398	0.419	-5.3#	82	0.00
46 T	Dibromomethane	0.275	0.302	-9.8	84	0.00
47 T	Bromodichloromethane	0.509	0.578	-13.6	86	0.00
48 T	Methyl methacrylate	0.504	0.533	-5.8	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
 Data File : VX039823.D
 Acq On : 18 Jan 2024 16:53
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 19 00:44:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	79	0.00
50 S	Toluene-d8	1.270	1.209	4.8	76	0.00
51 T	4-Methyl-2-Pentanone	0.707	0.732	-3.5	78	0.00
52 CM	Toluene	0.887	0.982	-10.7#	84	0.00
53 T	t-1,3-Dichloropropene	0.540	0.617	-14.3	83	0.00
54 T	cis-1,3-Dichloropropene	0.594	0.663	-11.6	82	0.00
55 T	1,1,2-Trichloroethane	0.377	0.411	-9.0	84	0.00
56 T	Ethyl methacrylate	0.611	0.668	-9.3	81	0.00
57 T	1,3-Dichloropropane	0.640	0.691	-8.0	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.289	0.294	-1.7	75	0.00
59 T	2-Hexanone	0.559	0.591	-5.7	78	0.00
60 T	Dibromochloromethane	0.356	0.422	-18.5	88	0.00
61 T	1,2-Dibromoethane	0.391	0.445	-13.8	86	0.00
62 S	4-Bromofluorobenzene	0.497	0.495	0.4	78	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.325	0.344	-5.8	87	0.00
65 PM	Chlorobenzene	1.041	1.120	-7.6	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.355	0.411	-15.8	87	0.00
67 C	Ethyl Benzene	1.933	2.103	-8.8#	84	0.00
68 T	m/p-Xylenes	0.711	0.771	-8.4	84	0.00
69 T	o-Xylene	0.704	0.749	-6.4	83	0.00
70 T	Styrene	1.150	1.262	-9.7	85	0.00
71 P	Bromoform	0.266	0.311	-16.9	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.981	4.074	-2.3	84	0.00
74 T	N-amyl acetate	2.075	2.164	-4.3	81	0.00
75 P	1,1,2,2-Tetrachloroethane	1.520	1.534	-0.9	83	0.00
76 T	1,2,3-Trichloropropane	1.296	1.287	0.7	83	0.00
77 T	Bromobenzene	0.886	0.914	-3.2	87	0.00
78 T	n-propylbenzene	4.818	5.053	-4.9	85	0.00
79 T	2-Chlorotoluene	2.978	2.985	-0.2	85	0.00
80 T	1,3,5-Trimethylbenzene	3.381	3.530	-4.4	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.427	0.436	-2.1	80	0.00
82 T	4-Chlorotoluene	3.378	3.473	-2.8	86	0.00
83 T	tert-Butylbenzene	3.228	3.343	-3.6	85	0.00
84 T	1,2,4-Trimethylbenzene	3.373	3.539	-4.9	85	0.00
85 T	sec-Butylbenzene	4.190	4.476	-6.8	87	0.00
86 T	p-Isopropyltoluene	3.349	3.587	-7.1	85	0.00
87 T	1,3-Dichlorobenzene	1.699	1.798	-5.8	89	0.00
88 T	1,4-Dichlorobenzene	1.755	1.831	-4.3	89	0.00
89 T	n-Butylbenzene	3.234	3.561	-10.1	87	0.00
90 T	Hexachloroethane	0.565	0.631	-11.7	87	0.00
91 T	1,2-Dichlorobenzene	1.680	1.770	-5.4	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.372	0.392	-5.4	83	0.00
93 T	1,2,4-Trichlorobenzene	1.026	1.093	-6.5	89	0.00
94 T	Hexachlorobutadiene	0.427	0.475	-11.2	91	0.00
95 T	Naphthalene	3.779	3.924	-3.8	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011824\
Data File : VX039823.D
Acq On : 18 Jan 2024 16:53
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 19 00:44:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.036	1.074	-3.7	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6