

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011921\Not Reviewed Data\
 Data File : VX020653.D
 Acq On : 19 Jan 2021 11:38
 Operator : JC/SP
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Jan 19 16:02:01 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 14:09:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	208116	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	337588	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	314998	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	153837	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	140073	49.23	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.46%
35) Dibromofluoromethane	5.465	113	113749	48.91	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.82%
50) Toluene-d8	8.696	98	415740	49.57	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.14%
62) 4-Bromofluorobenzene	11.122	95	157154	49.58	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.16%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.185	85	163042	51.45	ug/l 99
3) Chloromethane	1.307	50	148007	51.70	ug/l 100
4) Vinyl Chloride	1.392	62	155257	52.75	ug/l 99
5) Bromomethane	1.630	94	71470	52.20	ug/l 99
6) Chloroethane	1.709	64	94917	54.38	ug/l 99
7) Trichlorofluoromethane	1.916	101	219669	51.64	ug/l 98
8) Diethyl Ether	2.166	74	76218	50.84	ug/l 98
9) 1,1,2-Trichlorotrifluo...	2.368	101	117950	51.30	ug/l 99
10) Methyl Iodide	2.489	142	159940	51.94	ug/l 97
11) Tert butyl alcohol	3.038	59	120336	238.48	ug/l 99
12) 1,1-Dichloroethene	2.355	96	112299	50.38	ug/l 94
13) Acrolein	2.270	56	68857	226.83	ug/l 99
14) Allyl chloride	2.709	41	170171	50.64	ug/l 98
15) Acrylonitrile	3.111	53	320937	256.57	ug/l 98
16) Acetone	2.416	43	277447	251.07	ug/l 96
17) Carbon Disulfide	2.550	76	324962	51.00	ug/l 99
18) Methyl Acetate	2.745	43	126548	50.41	ug/l 100
19) Methyl tert-butyl Ether	3.166	73	382026	52.09	ug/l 100
20) Methylene Chloride	2.831	84	129803	48.39	ug/l 97
21) trans-1,2-Dichloroethene	3.142	96	124649	50.24	ug/l 93
22) Diisopropyl ether	3.825	45	353881	52.74	ug/l 94
23) Vinyl Acetate	3.782	43	1542487	269.41	ug/l 99
24) 1,1-Dichloroethane	3.666	63	223542	51.21	ug/l 99
25) 2-Butanone	4.629	43	420538	260.37	ug/l 99
26) 2,2-Dichloropropane	4.550	77	189494	51.03	ug/l 97
27) cis-1,2-Dichloroethene	4.562	96	140450	50.07	ug/l 100
28) Bromochloromethane	4.977	49	88905	48.76	ug/l 100
29) Tetrahydrofuran	5.087	42	264073	263.36	ug/l 99
30) Chloroform	5.172	83	233715	50.27	ug/l 98
31) Cyclohexane	5.544	56	192318	51.66	ug/l 98
32) 1,1,1-Trichloroethane	5.465	97	207400	50.72	ug/l 99
36) 1,1-Dichloropropene	5.769	75	173982	51.14	ug/l 100
37) Ethyl Acetate	4.788	43	163047	52.01	ug/l 99
38) Carbon Tetrachloride	5.757	117	185067	50.97	ug/l 97
39) Methylcyclohexane	7.440	83	205572	53.09	ug/l 98
40) Benzene	6.111	78	507863	50.94	ug/l 99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011921\Not Reviewed Data\
 Data File : VX020653.D
 Acq On : 19 Jan 2021 11:38
 Operator : JC/SP
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Jan 19 16:02:01 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 14:09:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	5.001	41	86274	51.04	ug/l	97
42) 1,2-Dichloroethane	6.160	62	189034	50.53	ug/l	98
43) Isopropyl Acetate	6.409	43	252155	52.09	ug/l	100
44) Trichloroethene	7.184	130	138099	50.89	ug/l	100
45) 1,2-Dichloropropane	7.489	63	128166	51.15	ug/l	99
46) Dibromomethane	7.635	93	92767	51.12	ug/l	96
47) Bromodichloromethane	7.873	83	185409	50.87	ug/l	98
48) Methyl methacrylate	7.745	41	127757	53.27	ug/l	98
49) 1,4-Dioxane	7.726	88	58880	1076.81	ug/l	97
51) 4-Methyl-2-Pentanone	8.616	43	810313	267.59	ug/l	99
52) Toluene	8.763	92	318116	51.91	ug/l	99
53) t-1,3-Dichloropropene	9.019	75	196466	52.72	ug/l	98
54) cis-1,3-Dichloropropene	8.415	75	216639	52.73	ug/l	97
55) 1,1,2-Trichloroethane	9.196	97	130772	50.45	ug/l	96
56) Ethyl methacrylate	9.159	69	184627	48.56	ug/l	99
57) 1,3-Dichloropropane	9.348	76	220777	51.60	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.287	63	463654	250.21	ug/l	98
59) 2-Hexanone	9.470	43	618500	272.66	ug/l	98
60) Dibromochloromethane	9.561	129	143369	51.30	ug/l	99
61) 1,2-Dibromoethane	9.653	107	138264	51.83	ug/l	99
64) Tetrachloroethene	9.317	164	123642	49.28	ug/l	93
65) Chlorobenzene	10.122	112	338361	49.70	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.201	131	128675	50.21	ug/l	99
67) Ethyl Benzene	10.232	91	614407	53.16	ug/l	99
68) m/p-Xylenes	10.342	106	458435	106.73	ug/l	97
69) o-Xylene	10.683	106	219391	53.19	ug/l	99
70) Styrene	10.695	104	374306	49.44	ug/l	97
71) Bromoform	10.842	173	101873	51.03	ug/l	98
73) Isopropylbenzene	11.000	105	593117	54.29	ug/l	99
74) N-amyl acetate	10.878	43	209385	48.80	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.250	83	194657	49.94	ug/l	98
76) 1,2,3-Trichloropropane	11.281	75	235335	63.75	ug/l	88
77) Bromobenzene	11.238	156	145633	52.11	ug/l	99
78) n-propylbenzene	11.342	91	686099	54.85	ug/l	99
79) 2-Chlorotoluene	11.402	91	413263	53.65	ug/l	99
80) 1,3,5-Trimethylbenzene	11.488	105	509269	55.76	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.055	75	65888	51.62	ug/l	99
82) 4-Chlorotoluene	11.494	91	485153	53.31	ug/l	100
83) tert-Butylbenzene	11.756	119	483002	54.44	ug/l	97
84) 1,2,4-Trimethylbenzene	11.793	105	510112	51.25	ug/l	100
85) sec-Butylbenzene	11.927	105	561491	54.65	ug/l	99
86) p-Isopropyltoluene	12.049	119	525267	50.76	ug/l	98
87) 1,3-Dichlorobenzene	12.006	146	271569	52.83	ug/l	98
88) 1,4-Dichlorobenzene	12.079	146	267926	51.49	ug/l	97
89) n-Butylbenzene	12.372	91	465400	50.51	ug/l	99
90) Hexachloroethane	12.579	117	91052	52.86	ug/l	91
91) 1,2-Dichlorobenzene	12.372	146	259698	51.80	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.981	75	45582	52.43	ug/l	92
93) 1,2,4-Trichlorobenzene	13.628	180	173013	49.18	ug/l	98
94) Hexachlorobutadiene	13.762	225	76023	49.86	ug/l	99
95) Naphthalene	13.817	128	583707	51.00	ug/l	99
96) 1,2,3-Trichlorobenzene	14.000	180	175383	55.31	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011921\Not Reviewed Data\
Data File : VX020653.D
Acq On : 19 Jan 2021 11:38
Operator : JC/SP
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Quant Time: Jan 19 16:02:01 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 14:09:08 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.629 min Scan# 745
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:168 Resp: 208116
Ion Ratio Lower Upper
168 100
99 60.3 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 51.45 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 85 Resp: 163042
Ion Ratio Lower Upper
85 100
87 32.2 16.4 49.2









#3
Chloromethane
Concen: 51.70 ug/l
RT: 1.307 min Scan# 36
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 50 Resp: 148007
Ion Ratio Lower Upper
50 100
52 33.2 26.4 39.6



#4
Vinyl Chloride
Concen: 52.75 ug/l
RT: 1.392 min Scan# 50
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 62 Resp: 155257
Ion Ratio Lower Upper
62 100
64 32.7 25.7 38.5









#5
Bromomethane
Concen: 52.20 ug/l
RT: 1.630 min Scan# 89
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 94 Resp: 71470
Ion Ratio Lower Upper
94 100
96 94.5 76.4 114.6



#6
Chloroethane
Concen: 54.38 ug/l
RT: 1.709 min Scan# 102
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 64 Resp: 94917
Ion Ratio Lower Upper
64 100
66 31.4 25.4 38.0









#7
Trichlorofluoromethane
Concen: 51.64 ug/l
RT: 1.916 min Scan# 136
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 219669
Ion Ratio Lower Upper
101 100
103 65.0 53.2 79.8



#8
Diethyl Ether
Concen: 50.84 ug/l
RT: 2.166 min Scan# 177
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 74 Resp: 76218
Ion Ratio Lower Upper
74 100
45 91.2 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.30 ug/l
RT: 2.368 min Scan# 210
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 117950
Ion Ratio Lower Upper
101 100
85 42.6 34.9 52.3
151 77.0 61.4 92.2



#10
Methyl Iodide
Concen: 51.94 ug/l
RT: 2.489 min Scan# 230
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:142 Resp: 159940
Ion Ratio Lower Upper
142 100
127 44.5 33.3 49.9
141 14.1 11.4 17.0









#11

Tert butyl alcohol
Concen: 238.48 ug/l
RT: 3.038 min Scan# 320
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 59 Resp: 120336
Ion Ratio Lower Upper
59 100
57 10.1 7.8 11.8




#12

1,1-Dichloroethene
Concen: 50.38 ug/l
RT: 2.355 min Scan# 208
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 96 Resp: 112299
Ion Ratio Lower Upper
96 100
61 164.1 124.6 187.0
98 66.0 50.6 76.0









#13
Acrolein
Concen: 226.83 ug/l
RT: 2.270 min Scan# 194
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 68857
Ion Ratio Lower Upper
56 100
55 69.3 56.4 84.6



#14
Allyl chloride
Concen: 50.64 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 41 Resp: 170171
Ion Ratio Lower Upper
41 100
39 66.0 51.6 77.4
76 39.5 30.5 45.7









#15
Acrylonitrile
Concen: 256.57 ug/l
RT: 3.111 min Scan# 332
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 53 Resp: 320937




Ion	Ratio	Lower	Upper
53	100		
52	83.4	65.3	97.9
51	36.5	28.2	42.2




#16
Acetone
Concen: 251.07 ug/l
RT: 2.416 min Scan# 218
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 43 Resp: 277447

Ion	Ratio	Lower	Upper
43	100		
58	32.7	27.8	41.8









#17
Carbon Disulfide
Concen: 51.00 ug/l
RT: 2.550 min Scan# 240
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 76 Resp: 324962
Ion Ratio Lower Upper
76 100
78 9.4 7.2 10.8



#18
Methyl Acetate
Concen: 50.41 ug/l
RT: 2.745 min Scan# 272
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 43 Resp: 126548
Ion Ratio Lower Upper
43 100
74 27.4 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 52.09 ug/l
RT: 3.166 min Scan# 341
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 73 Resp: 382026
Ion Ratio Lower Upper
73 100
57 21.3 17.1 25.7



#20
Methylene Chloride
Concen: 48.39 ug/l
RT: 2.831 min Scan# 286
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 84 Resp: 129803
Ion Ratio Lower Upper
84 100
49 113.8 88.6 132.8
51 36.1 27.0 40.6
86 64.6 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 50.24 ug/l
RT: 3.142 min Scan# 337
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 96 Resp: 124649
Ion Ratio Lower Upper
96 100
61 144.5 107.7 161.5
98 66.4 50.4 75.6



#22
Diisopropyl ether
Concen: 52.74 ug/l
RT: 3.825 min Scan# 449
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 45 Resp: 353881
Ion Ratio Lower Upper
45 100
43 64.4 58.0 87.0
87 31.8 25.4 38.0
59 13.1 10.2 15.2









#23
 Vinyl Acetate
 Concen: 269.41 ug/l
 RT: 3.782 min Scan# 442
 Delta R.T. -0.000 min
 Lab File: VX020653.D
 Acq: 19 Jan 2021 11:38




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 43 Resp: 1542487
 Ion Ratio Lower Upper
 43 100
 86 12.1 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 51.21 ug/l
 RT: 3.666 min Scan# 423
 Delta R.T. 0.000 min
 Lab File: VX020653.D
 Acq: 19 Jan 2021 11:38

Tgt Ion: 63 Resp: 223542
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.8 11.2
 100 4.8 2.5 7.5









#25
2-Butanone
Concen: 260.37 ug/l
RT: 4.629 min Scan# 581
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 420538
Ion Ratio Lower Upper
43 100
72 30.1 23.7 35.5



#26
2,2-Dichloropropane
Concen: 51.03 ug/l
RT: 4.550 min Scan# 568
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 77 Resp: 189494
Ion Ratio Lower Upper
77 100
97 21.9 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 50.07 ug/l

RT: 4.562 min Scan# 570





Delta R.T. 0.000 min

Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Tgt Ion: 96 Resp: 140450

Ion	Ratio	Lower	Upper
96	100		
61	142.9	0.0	285.6
98	64.6	0.0	129.6



#28

Bromochloromethane

Concen: 48.76 ug/l

RT: 4.977 min Scan# 638




Delta R.T. -0.000 min


Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Tgt Ion: 49 Resp: 88905

Ion	Ratio	Lower	Upper
49	100		
129	2.3	0.0	4.0
130	85.1	68.2	102.2





#29

Tetrahydrofuran

Concen: 263.36 ug/l

RT: 5.087 min Scan# 656

Delta R.T. -0.000 min





Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 42 Resp: 264073

Ion	Ratio	Lower	Upper
42	100		
72	52.7	41.7	62.5
71	48.7	38.4	57.6



#30

Chloroform

Concen: 50.27 ug/l

RT: 5.172 min Scan# 670




Delta R.T. -0.000 min


Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Tgt Ion: 83 Resp: 233715

Ion	Ratio	Lower	Upper
83	100		
85	66.6	52.0	78.0









#31
Cyclohexane
Concen: 51.66 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 56 Resp: 192318




Ion	Ratio	Lower	Upper
56	100		
69	33.0	25.3	37.9
84	91.2	71.4	107.0




#32
1,1,1-Trichloroethane
Concen: 50.72 ug/l
RT: 5.465 min Scan# 718
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 97 Resp: 207400

Ion	Ratio	Lower	Upper
97	100		
99	63.8	51.5	77.3
61	42.9	34.2	51.4









#33

1,2-Dichloroethane-d4
Concen: 49.23 ug/l
RT: 6.025 min Scan# 810
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 65 Resp: 140073
Ion Ratio Lower Upper
65 100
67 51.9 0.0 105.4




#34

1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.830 min Scan# 942
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 114 Resp: 337588
Ion Ratio Lower Upper
114 100
63 19.7 0.0 40.8
88 16.6 0.0 32.8





#35

Dibromofluoromethane

Concen: 48.91 ug/l

RT: 5.465 min Scan# 718

Delta R.T. -0.000 min

Lab File: VX020653.D

Acq: 19 Jan 2021 11:38





Tgt Ion:113 Resp: 113749

Ion Ratio Lower Upper

113 100

111 103.1 81.9 122.9

192 19.3 16.7 25.1



#36

1,1-Dichloropropene

Concen: 51.14 ug/l

RT: 5.769 min Scan# 768

Delta R.T. 0.000 min

Lab File: VX020653.D

Acq: 19 Jan 2021 11:38




Tgt Ion: 75 Resp: 173982


Ion Ratio Lower Upper

75 100

110 35.8 18.1 54.4

77 31.6 25.3 37.9





#37

Ethyl Acetate

Concen: 52.01 ug/l

RT: 4.788 min Scan# 607





Delta R.T. -0.000 min

Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Tgt Ion: 43 Resp: 163047

Ion	Ratio	Lower	Upper
43	100		
61	15.2	12.8	19.2
70	12.8	10.5	15.7



#38

Carbon Tetrachloride

Concen: 50.97 ug/l

RT: 5.757 min Scan# 766




Delta R.T. 0.000 min


Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Tgt Ion: 117 Resp: 185067

Ion	Ratio	Lower	Upper
117	100		
119	92.7	76.2	114.2
121	29.7	24.8	37.2





#39

Methylcyclohexane

Concen: 53.09 ug/l

RT: 7.440 min Scan# 1042

Delta R.T. -0.000 min

Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 83 Resp: 205572

Ion Ratio Lower Upper

83 100

55 72.1 58.9 88.3

98 49.4 38.1 57.1



#40

Benzene

Concen: 50.94 ug/l

RT: 6.111 min Scan# 824

Delta R.T. -0.000 min

Lab File: VX020653.D




Acq: 19 Jan 2021 11:38


Tgt Ion: 78 Resp: 507863

Ion Ratio Lower Upper

78 100

77 24.0 18.9 28.3









#41
Methacrylonitrile
Concen: 51.04 ug/l
RT: 5.001 min Scan# 642
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 41 Resp: 86274




Ion	Ratio	Lower	Upper
41	100		
39	54.8	42.5	63.7
67	84.8	65.0	97.6
52	33.3	25.3	37.9




#42
1,2-Dichloroethane
Concen: 50.53 ug/l
RT: 6.160 min Scan# 832
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 62 Resp: 189034

Ion	Ratio	Lower	Upper
62	100		
98	9.4	0.0	20.2









#43
Isopropyl Acetate
Concen: 52.09 ug/l
RT: 6.409 min Scan# 873
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 252155
Ion Ratio Lower Upper
43 100
61 22.6 18.2 27.4
87 15.7 12.5 18.7



#44
Trichloroethene
Concen: 50.89 ug/l
RT: 7.184 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:130 Resp: 138099
Ion Ratio Lower Upper
130 100
95 98.8 0.0 198.0





#45

1,2-Dichloropropane

Concen: 51.15 ug/l

RT: 7.489 min Scan# 1050

Delta R.T. -0.000 min

Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Tgt Ion: 63 Resp: 128166

Ion Ratio Lower Upper

63 100





65 31.7 25.7 38.5

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050



#46

Dibromomethane

Concen: 51.12 ug/l

RT: 7.635 min Scan# 1074

Delta R.T. -0.000 min

Lab File: VX020653.D

Acq: 19 Jan 2021 11:38




Tgt Ion: 93 Resp: 92767


Ion Ratio Lower Upper

93 100

95 82.5 66.8 100.2

174 98.0 83.1 124.7









#47
Bromodichloromethane
Concen: 50.87 ug/l
RT: 7.873 min Scan# 1113
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 185409
Ion Ratio Lower Upper
83 100
85 65.8 51.4 77.0
127 9.0 6.8 10.2



#48
Methyl methacrylate
Concen: 53.27 ug/l
RT: 7.745 min Scan# 1092
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 41 Resp: 127757
Ion Ratio Lower Upper
41 100
69 96.7 76.0 114.0
39 56.7 44.6 67.0





#49

1,4-Dioxane

Concen: 1076.81 ug/l

RT: 7.726 min Scan# 1089





Delta R.T. 0.000 min

Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Tgt Ion: 88 Resp: 58880

Ion	Ratio	Lower	Upper
88	100		
43	30.2	23.2	34.8
58	65.7	55.0	82.4



#50

Toluene-d8

Concen: 49.57 ug/l

RT: 8.696 min Scan# 1248




Delta R.T. -0.000 min


Lab File: VX020653.D

Acq: 19 Jan 2021 11:38

Tgt Ion: 98 Resp: 415740

Ion	Ratio	Lower	Upper
98	100		
100	65.0	51.9	77.9









#51
4-Methyl-2-Pentanone
Concen: 267.59 ug/l
RT: 8.616 min Scan# 1235
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 810313
Ion Ratio Lower Upper
43 100
58 43.9 35.4 53.0



#52
Toluene
Concen: 51.91 ug/l
RT: 8.763 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 92 Resp: 318116
Ion Ratio Lower Upper
92 100
91 172.5 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 52.72 ug/l
RT: 9.019 min Scan# 1301
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 75 Resp: 196466
Ion Ratio Lower Upper
75 100
77 32.1 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 52.73 ug/l
RT: 8.415 min Scan# 1202
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 75 Resp: 216639
Ion Ratio Lower Upper
75 100
77 32.0 25.7 38.5
39 40.8 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 50.45 ug/l
RT: 9.196 min Scan# 1330
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 97 Resp: 130772




Ion	Ratio	Lower	Upper
97	100		
83	82.2	69.3	103.9
85	52.5	44.3	66.5
99	61.3	50.3	75.5




#56
Ethyl methacrylate
Concen: 48.56 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 69 Resp: 184627





Ion	Ratio	Lower	Upper
69	100		
41	57.2	46.8	70.2
39	31.2	24.7	37.1








#57
1,3-Dichloropropane
Concen: 51.60 ug/l
RT: 9.348 min Scan# 1355
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 76 Resp: 220777
Ion Ratio Lower Upper
76 100
78 32.3 26.2 39.2




#58
2-Chloroethyl Vinyl ether
Concen: 250.21 ug/l
RT: 8.287 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 63 Resp: 463654
Ion Ratio Lower Upper
63 100
106 30.3 23.5 35.3







Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



#59
2-Hexanone
Concen: 272.66 ug/l
RT: 9.470 min Scan# 1375
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 618500
Ion Ratio Lower Upper
43 100
58 60.5 30.9 92.8



#60
Dibromochloromethane
Concen: 51.30 ug/l
RT: 9.561 min Scan# 1390
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 129 Resp: 143369
Ion Ratio Lower Upper
129 100
127 76.3 38.4 115.2









#61
1,2-Dibromoethane
Concen: 51.83 ug/l
RT: 9.653 min Scan# 1405
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:107 Resp: 138264
Ion Ratio Lower Upper
107 100
109 96.0 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 49.58 ug/l
RT: 11.122 min Scan# 1646
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 95 Resp: 157154
Ion Ratio Lower Upper
95 100
174 77.1 0.0 154.6
176 72.4 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.092 min Scan# 1477
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:117 Resp: 314998




Ion	Ratio	Lower	Upper
117	100		
82	56.7	47.4	71.2
119	32.1	26.6	39.8




#64
Tetrachloroethene
Concen: 49.28 ug/l
RT: 9.317 min Scan# 1350
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:164 Resp: 123642

Ion	Ratio	Lower	Upper
164	100		
166	133.9	103.4	155.2
129	100.7	70.8	106.2
131	94.4	72.5	108.7









#65
Chlorobenzene
Concen: 49.70 ug/l
RT: 10.122 min Scan# 1482
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:112 Resp: 338361
Ion Ratio Lower Upper
112 100
114 33.1 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 50.21 ug/l
RT: 10.201 min Scan# 1495
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:131 Resp: 128675
Ion Ratio Lower Upper
131 100
133 96.5 47.7 143.1
119 64.0 32.1 96.5









#67
Ethyl Benzene
Concen: 53.16 ug/l
RT: 10.232 min Scan# 1500
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 614407
Ion Ratio Lower Upper
91 100
106 31.1 24.6 37.0



#68
m/p-Xylenes
Concen: 106.73 ug/l
RT: 10.342 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 106 Resp: 458435
Ion Ratio Lower Upper
106 100
91 208.6 162.8 244.2









#69
o-Xylene
Concen: 53.19 ug/l
RT: 10.683 min Scan# 1574
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:106 Resp: 219391
Ion Ratio Lower Upper
106 100
91 215.2 108.1 324.2



#70
Styrene
Concen: 49.44 ug/l
RT: 10.695 min Scan# 1576
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:104 Resp: 374306
Ion Ratio Lower Upper
104 100
78 52.2 40.0 60.0
103 55.8 43.4 65.0









#71
Bromoform
Concen: 51.03 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:173 Resp: 101873
Ion Ratio Lower Upper
173 100
175 49.5 23.9 71.8
254 0.1 0.1 0.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.061 min Scan# 1800
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:152 Resp: 153837
Ion Ratio Lower Upper
152 100
115 88.3 41.1 123.3
150 177.3 0.0 349.0









#73
Isopropylbenzene
Concen: 54.29 ug/l
RT: 11.000 min Scan# 1626
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:105 Resp: 593117
Ion Ratio Lower Upper
105 100
120 26.0 13.2 39.6



#74
N-ethyl acetate
Concen: 48.80 ug/l
RT: 10.878 min Scan# 1606
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 43 Resp: 209385
Ion Ratio Lower Upper
43 100
70 51.3 39.0 58.6
55 28.7 21.5 32.3
61 27.4 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 49.94 ug/l
RT: 11.250 min Scan# 1667
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 194657
Ion Ratio Lower Upper
83 100
131 10.6 5.1 15.2
85 65.2 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 63.75 ug/l
RT: 11.281 min Scan# 1672
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 75 Resp: 235335
Ion Ratio Lower Upper
75 100
77 31.0 19.1 57.5









#77
Bromobenzene
Concen: 52.11 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:156 Resp: 145633
Ion Ratio Lower Upper
156 100
77 166.1 82.8 248.6
158 97.4 49.1 147.4



#78
n-propylbenzene
Concen: 54.85 ug/l
RT: 11.342 min Scan# 1682
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 91 Resp: 686099
Ion Ratio Lower Upper
91 100
120 22.7 11.5 34.5









#79
2-Chlorotoluene
Concen: 53.65 ug/l
RT: 11.402 min Scan# 1692
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 413263
Ion Ratio Lower Upper
91 100
126 32.9 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 55.76 ug/l
RT: 11.488 min Scan# 1706
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 105 Resp: 509269
Ion Ratio Lower Upper
105 100
120 48.3 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 51.62 ug/l
RT: 11.055 min Scan# 1635
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 75 Resp: 65888
Ion Ratio Lower Upper
75 100
53 94.5 74.5 111.7
89 46.6 37.9 56.9



#82
4-Chlorotoluene
Concen: 53.31 ug/l
RT: 11.494 min Scan# 1707
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 91 Resp: 485153
Ion Ratio Lower Upper
91 100
126 29.5 14.9 44.5









#83
tert-Butylbenzene
Concen: 54.44 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:119 Resp: 483002
Ion Ratio Lower Upper
119 100
91 58.6 27.9 83.7
134 22.7 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 51.25 ug/l
RT: 11.793 min Scan# 1756
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:105 Resp: 510112
Ion Ratio Lower Upper
105 100
120 45.2 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 54.65 ug/l
 RT: 11.927 min Scan# 1778
 Delta R.T. -0.000 min
 Lab File: VX020653.D
 Acq: 19 Jan 2021 11:38




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:105 Resp: 561491
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 50.76 ug/l
 RT: 12.049 min Scan# 1798
 Delta R.T. -0.000 min
 Lab File: VX020653.D
 Acq: 19 Jan 2021 11:38

Tgt Ion:119 Resp: 525267
 Ion Ratio Lower Upper
 119 100
 134 26.5 12.9 38.7
 91 24.7 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 52.83 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampled :
VSTDICCC050


Tgt Ion:146 Resp: 271569
Ion Ratio Lower Upper
146 100
111 40.8 20.2 60.5
148 61.8 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 51.49 ug/l
RT: 12.079 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:146 Resp: 267926
Ion Ratio Lower Upper
146 100
111 40.8 19.6 58.8
148 65.7 31.6 95.0









#89
n-Butylbenzene
Concen: 50.51 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 465400
Ion Ratio Lower Upper
91 100
92 54.6 27.4 82.0
134 25.4 13.2 39.5



#90
Hexachloroethane
Concen: 52.86 ug/l
RT: 12.579 min Scan# 1885
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 117 Resp: 91052
Ion Ratio Lower Upper
117 100
201 73.2 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 51.80 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 259698
Ion Ratio Lower Upper
146 100
111 43.4 20.8 62.2
148 63.4 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 52.43 ug/l
RT: 12.981 min Scan# 1951
Delta R.T. 0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion: 75 Resp: 45582
Ion Ratio Lower Upper
75 100
155 83.3 45.3 135.9
157 106.6 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 49.18 ug/l
RT: 13.628 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:	180	Resp:	173013
Ion Ratio	Lower	Upper	
180	100		
182	94.3	47.7	143.1
145	31.5	15.0	45.0



#94
Hexachlorobutadiene
Concen: 49.86 ug/l
RT: 13.762 min Scan# 2079
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:	225	Resp:	76023
Ion Ratio	Lower	Upper	
225	100		
223	62.1	31.1	93.2
227	61.6	31.6	95.0









#95
Naphthalene
Concen: 51.00 ug/l
RT: 13.817 min Scan# 2088
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:128 Resp: 583707
Ion Ratio Lower Upper
128 100
127 12.4 10.3 15.5
129 10.8 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 55.31 ug/l
RT: 14.000 min Scan# 2118
Delta R.T. -0.000 min
Lab File: VX020653.D
Acq: 19 Jan 2021 11:38

Tgt Ion:180 Resp: 175383
Ion Ratio Lower Upper
180 100
182 92.3 47.6 142.9
145 33.1 16.1 48.3

