

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011921\
 Data File : VX020658.D
 Acq On : 19 Jan 2021 13:52
 Operator : JC/SP
 Sample : VX0119MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0119MBL01

Quant Time: Jan 20 03:19:51 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 14:09:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	234397	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	400552	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	371848	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	168906	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	160140	49.97	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.94%
35) Dibromofluoromethane	5.458	113	134707	48.81	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.62%
50) Toluene-d8	8.696	98	493317	49.57	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.14%
62) 4-Bromofluorobenzene	11.116	95	179359	47.69	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.38%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011921\
Data File : VX020658.D
Acq On : 19 Jan 2021 13:52
Operator : JC/SP
Sample : VX0119MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0119MBL01

Quant Time: Jan 20 03:19:51 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 14:09:08 2021
Response via : Initial Calibration

