

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011921\
 Data File : VX020667.D
 Acq On : 19 Jan 2021 18:26
 Operator : JC/SP
 Sample : M1158-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 20 03:21:47 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 14:09:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	239575	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	416719	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	394930	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	172969	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	165018	50.38	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.76%
35) Dibromofluoromethane	5.464	113	137763	47.98	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.96%
50) Toluene-d8	8.695	98	507516	49.02	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.04%
62) 4-Bromofluorobenzene	11.122	95	185010	47.28	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.56%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011921\
Data File : VX020667.D
Acq On : 19 Jan 2021 18:26
Operator : JC/SP
Sample : M1158-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 20 03:21:47 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 14:09:08 2021
Response via : Initial Calibration

