

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011921\
 Data File : VX020668.D
 Acq On : 19 Jan 2021 18:49
 Operator : JC/SP
 Sample : M1158-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 20 03:21:57 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 14:09:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	240285	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	412945	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	388390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	170187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	166840	50.78	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.56%
35) Dibromofluoromethane	5.464	113	136833	48.09	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.18%
50) Toluene-d8	8.696	98	507240	49.44	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.88%
62) 4-Bromofluorobenzene	11.122	95	183017	47.20	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011921\
Data File : VX020668.D
Acq On : 19 Jan 2021 18:49
Operator : JC/SP
Sample : M1158-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 20 03:21:57 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 14:09:08 2021
Response via : Initial Calibration

