

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
 Data File : VX026512.D
 Acq On : 19 Jan 2022 12:28
 Operator : JC/MD
 Sample : N1187-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20904

Quant Time: Jan 20 05:04:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	146631	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	245592	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209102	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87879	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97073	47.391	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.780%
35) Dibromofluoromethane	5.391	113	77741	49.018	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.040%
50) Toluene-d8	8.653	98	281155	49.195	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.400%
62) 4-Bromofluorobenzene	11.079	95	99043	48.784	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.560%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	7893	7.556	ug/l	95
40) Benzene	6.043	78	24212	3.499	ug/l	97
52) Toluene	8.720	92	14245	3.357	ug/l	99
67) Ethyl Benzene	10.195	91	2694	0.337	ug/l	94
68) m/p-Xylenes	10.305	106	6258	2.090	ug/l	96
69) o-Xylene	10.640	106	5894	2.040	ug/l	100
70) Styrene	10.658	104	4979	1.056	ug/l	92
80) 1,3,5-Trimethylbenzene	11.451	105	1491	0.251	ug/l	94
84) 1,2,4-Trimethylbenzene	11.756	105	7149	1.204	ug/l	90
95) Naphthalene	13.780	128	83676	13.386	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
 Data File : VX026512.D
 Acq On : 19 Jan 2022 12:28
 Operator : JC/MD
 Sample : N1187-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20904

Quant Time: Jan 20 05:04:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

