

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
 Data File : VX026513.D
 Acq On : 19 Jan 2022 12:51
 Operator : JC/MD
 Sample : N1187-02 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20963

Quant Time: Jan 20 05:04:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141680	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240142	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85686	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96545	48.781	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.560%
35) Dibromofluoromethane	5.391	113	76311	49.209	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.420%
50) Toluene-d8	8.653	98	275238	49.253	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.500%
62) 4-Bromofluorobenzene	11.079	95	96275	48.497	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.000%
Target Compounds						
16) Acetone	2.392	43	2958	2.931	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
Data File : VX026513.D
Acq On : 19 Jan 2022 12:51
Operator : JC/MD
Sample : N1187-02 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20963

Quant Time: Jan 20 05:04:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

