

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
 Data File : VX026514.D
 Acq On : 19 Jan 2022 13:14
 Operator : JC/MD
 Sample : N1187-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20988

Quant Time: Jan 20 05:04:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	141129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	237509	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87988	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95941	48.665	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.320%
35) Dibromofluoromethane	5.391	113	74826	48.786	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.580%
50) Toluene-d8	8.653	98	277059	50.128	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.260%
62) 4-Bromofluorobenzene	11.079	95	100468	51.170	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.340%
Target Compounds						
16) Acetone	2.392	43	1676	1.667	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
Data File : VX026514.D
Acq On : 19 Jan 2022 13:14
Operator : JC/MD
Sample : N1187-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20988

Quant Time: Jan 20 05:04:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

